VOLATILE ORGANIC COMPOUND EMISSION TEST REPORT

OF THE

CHEVRON PRODUCTS COMPANY

PORT EVERGLADES, FLORIDA TRANSPORT LOADING TERMINAL

ON THE

JOHN ZINK CARBON VAPOR RECOVERY UNIT

ON

JULY 14, 2011

Terminal Info:

Chevron Product Company

1400 SE 24th Street

Ft. Lauderdale, Florida 33316

Phone:954-764-2309 Fax: 954-764-2533

Test Performed by:

Tony Fenton

Jordan Technologies, Inc.

2820 South English Station Rd.

Louisville, KY 40299 (502) 267- 8344

E-mail tfenton@jordantech.com



In reference to the Chevron Products Air Emission Source Test conducted at the Port Everglades, Florida Transport Loading Facility on July 14, 2011 and described in the following report;

I certify under penalty of law that the information provided in this document is true, accurate and complete. I am aware that there are significant civil and criminal penalties, including fines or imprisonment or both, for submitting false, inaccurate or incomplete information.

Tony Fenton

Technical Service Group Jordan Technologies

TABLE OF CONTENTS

EXECUTIVE SUMMARY1
TERMINAL OPERATION AND DESCRIPTION2
VAPOR RECOVERY SYSTEM3
MEASUREMENT TECHNIQUES4
TEST EQUIPMENT5
EXAMPLE CALCULATIONS6-7
TEST RESULTS8
VAPOR RECOVERY UNIT9
VAPOR RECOVERY UNIT TEST SCHEMATIC10
APPENDIX ATRUCK MONITORING DATA SHEETS
APPENDIX BINSTRUMENT AND CALIBRATION INFORMATION
APPENDIX CCOMPUTER PRINTOUT, VOC ANALYZER STRIP CHARTS

EXECUTIVE SUMMARY

The Chevron Products terminal in Port Everglades, Florida is a bulk transport loading facility for Gasoline Products.

The products are bottom loaded into transport tankers and the displaced hydrocarbon vapors are balanced to a JOHN ZINK CARBON ADSORPTION / ABSORPTION VAPOR RECOVERY UNIT (VRU).

This facility was source tested for air emissions on July 14, 2011. The purpose of this test was to confirm proper operation of the VRU and verify compliance with applicable VOC (Volatile Organic Compound) air emission requirements.

The Gasoline Terminal Air Emission Source Test was conducted in accordance with procedures established, and the test methods referenced, in the Code of Federal Regulations; CFR 40, Part 60, Subpart XX & CFR 40, Part 63, Subpart BBBBBB. Specific procedures used include:

EPA TEST METHOD

<u>MEASUREMENT</u>

Method 2A Method 25B Method 21 40 CFR 60 Subsection 60.503 (d) 40 CFR 63 Subpart BBBBBB Exhaust Vapor Volume
Inlet and Outlet VOC Concentrations
Potential Leak Sources
Transport Loading Maximum Backpressure
GACT Rule Leak Rate (500 PPM)

The results of this air emission test demonstrate that this source is in compliance with all applicable Federal and Local requirements. A summary of the data is presented below:

TEST PARAMETER

MEASURED VALUE

REQUIRED VALUE

VOC Emissions

0.28 mg/liter

12 mg/liter

Method 21 Leak Testing was performed on the day prior to testing. A portable LEL Meter was calibrated using a 500 PPM Methane calibration gas and used to check for Leaks around all fittings, flanges, valves and any other exposed potential leak source. No Leaks above 500 PPM were found.

TERMINAL OPERATION AND DESCRIPTION

Light petroleum products are bottom loaded at five loading racks at the Chevron Products Company, Port Everglades, Florida Terminal.

The terminal is equipped to load Regular, Midgrade, and Premium Unleaded Gasoline onto transports. Diesel Fuel, Jet Fuel, AV Gas and Transmix are also available for loading onto transports.

Each rack is equipped with vapor recovery hoses positioned at the transport loading positions for hook up to the Vapor Control System. The vapor hoses and associated piping transports the vapors to the VRU. The system also employs a liquid knock-out tank and pressure / vacuum relief vent upstream from the VRU.

NOTE: The amount of product loaded during the test at the East rack was obtained by taking the meter readings before and after the test. Total Diesel Gallons loaded 55,208. These gallons were added to the West Rack gallons in order to obtain total gallons loaded during the test. This rack loads Diesel and Jet only.

<u>JOHN ZINK VAPOR RECOVERY UNIT</u>

The terminal is equipped with a John Zink Adsorption / Absorption Gasoline Vapor Recovery Unit. Hydrocarbon vapors enter the John Zink VRU into one of two Carbon Adsorbers. The Hydrocarbon - air mixture flows up through the adsorber where the bulk of the hydrocarbons are adsorbed. The air continues through the Carbon Absorber and is vented to the atmosphere. The saturated carbon is then desorbed by employing vacuum regeneration at 27.5" Hg Vacuum, while the second Carbon Absorber is receiving the hydrocarbon - air mixture generated in transport loading activity. The purpose of regeneration is to restore the carbon to a level where it will effectively absorb hydrocarbons again. The two carbon adsorbers alternate between adsorption and regeneration at 15 minute intervals.

When a Carbon Absorber is in the regeneration mode, a liquid ring vacuum pump pulls the hydrocarbon from the carbon. The rich hydrocarbon vapors from the Carbon Absorber are mixed with the vacuum pump seal fluid and are discharged to an Absorber / Separator.

The liquid hydrocarbons are condensed and separated from the seal fluid in the separator compartment and are discharged back to a holding tank. Any remaining hydrocarbons pass up through the packed Absorber tower and are contacted by a fresh stream of gasoline which absorbs most of the remaining hydrocarbons. The small amount of hydrocarbons that is left then leaves the top of the absorber and is directed back to the Carbon Adsorber where the whole process is repeated.

The John Zink Vapor Recovery Unit is illustrated schematically on page 9.

MEASUREMENT AND DATA ANALYSIS

The NonDispersive InfraRed (NDIR) analyzer, turbine flow meter, exhaust vapor thermistor and exhaust pressure transducer are connected to the VRU exhaust stack in order to acquire their respective data. A quad check valve assembly is employed to provide for proper VRU regeneration air flow and allow one turbine meter to satisfy both carbon vessel measurement requirements.

The barometric pressure transducer and ambient thermistor are located in close proximity to the VRU in order to acquire ambient atmospheric conditions for use in subsequent standardization equations. A test schematic depicting general test equipment configuration is included on page 10.

Each transducer data channel is scaled and connected to the computer input board. Using an operations code program each input channel is read 25 times in a 5 second interval and mass, flow, concentration, temperature, and pressure values are averaged and stored in an array for subsequent use.

After sixty 5 second intervals (5 minutes) the hard disk array is polled and average values are determined for concentration, pressure, and temperature. These values along with the flow for the 5 minute period are used to compute the mass emitted for that 5 minute period. These averaged and summed values are then printed out as the 5 minute interval data and are again stored on hard disk until the six hour test period is completed.

Upon completion of the test, the 5 minute interval data is polled to determine test averages for Inlet and Outlet VOC concentration, pressure and temperature data for all test intervals during which VRU exhaust flow was greater than zero and volume and milligram emission data is summed for all 5 minute periods to arrive at a final test period total.

This data acquisition methodology essentially represents a series of very short (5 second) intervals during which VRU operation is measured, averaged and standardized. This effectively removes all judgmental decisions from data reduction processes and provides a technically unbiased analysis of VRU operation.

Additionally, pretest and post test vapor analyzer calibrations are conducted, along with an hourly analyzer calibration drift check verification. Following the conclusion of the six hour test the loading rack volumes are calculated and final mass emission values are determined. Copies of the transport loading rack sheets, hydrocarbon analyzer strip charts and computer print outs are attached as Appendices to this test report.

TEST EQUIPMENT

Quantity	Item
2	Thermistor Temperature Probes
1	IBM Compatible Computer with 16 Channel, 12 bit A/D Input Card
1	RKI Eagle Instrument Combustible Gas Indicator
1	Setra Model #261 (or #264) Variable Differential Pressure Transducer
1	Setraceram Model #361 (or #304) Digital Barometer
1	American Meter Co. 8" Turbine Flow Meter
1 (or 2)	Strip Chart Recorder, either: Yokogawa DX 1000n Paperless Recorder
2	NonDispersive InfraRed Analyzers (NDIR), either: Horiba VIA-510 Enviromax 2010 NDIR Analyzer

All equipment specifications are shown in Appendix B along with available calibration and accuracy information.

EXAMPLE CALCULATIONS

A. Terminology:

T_a = Ambient Temperature (° Celsius).

P_b = Barometric Pressure (mm Hg).

L_t = Total volume of liquid dispensed from all controlled racks during the test period (Liters).

V_e = Volume of air-hydrocarbon mixture exhausted from the processing unit (cubic meters).

V_{es} = Normalized volume of air-hydrocarbon mixture exhausted (Cubic meters at 20° Celsius, 760 mm Hg).

C_e = Volume fraction of hydrocarbons in exhausted mixture (Volume % as C₃H₈/100, corrected for methane content, if required).

T_e = Temperature at process unit exhaust (° Celsius).

P_e = Pressure at processing unit exhaust (mm Hg. absolute).

M_e = Mass of VOC emitted (milligrams).

(M/L)_e = Mass of hydrocarbons exhausted from the processing unit per volume of liquid loaded (mg/liter).

 $(M/T)_e$ = Mass of hydrocarbons exhausted from the processing unit per unit time (lb/hour).

Constants:

 $0.3858 = (273.2^{\circ} \text{ C} + 20^{\circ} \text{ C}) / (760 \text{ mm Hg})$ Normalization Factor.

 $1.83 \times 10^6 \text{ mg/m}^3 = \text{Standard Density of Propane } (C^3 H^8).$

454,000 = Conversion Factor mg/lb.

3.785 = Conversion Factor Liter/Gallon.

264.2 = Conversion Factor gallons / meter³

B. <u>Calculate the Following Results for Each Period of the Vapor Control System Operation:</u>

(1.) Volume of air-hydrocarbon mixture exhausted from the vapor control system:

 $V_e = (V_{ef} - V_{ei})$ (meters³)

(where subscript f refers to final and subscript refers to initial)

 V_e = Totalized volume from flow rate and time records.

(2.) Normalized volume of exhausted mixture:

 $V_{es} = (0.3858^{\circ} \text{ Kelvin/mm Hg}) \times V_{e} \times P_{e}$ (meter³) (T_e + 273.2)

(3.) Mass of hydrocarbons exhausted from the vapor control system:

$$M_e = (1.83 \times 10^6 \text{ mg C}^3 \text{H}^8) \times (V_{es}) \times (C_e)$$
 (mg) (equation B)
meter3

C. Calculate the Average Mass of Hydrocarbons Emitted Per Volume of Gasoline Loaded:

$$(M/L)_e = M_e/L_t$$

(mg/liter)

D. Calculate the Average Mass of Hydrocarbons Emitted Per Unit Time:

$$(M/T)_e = (M/L)_e \times \frac{1}{454,000 \text{ mg}} \frac{1 \text{ lb} \times 3.785 \text{ liter}}{1 \text{ gal}} \times \frac{\text{Acct. Gal}}{1 \text{ Test Time}}$$
 (lb/hr)

E. Calculation for Efficiency (if used):

Unit Efficiency = [1-(outlet mg / inlet mg)] x 100%

Where inlet milligrams is derived using inlet concentration and volume of liquid loaded onto transports, assuming a vapor growth ratio of 1:1 and no gross leaks.

F. Example ME Calculation For a Typical Five Minute Interval:

This is an example calculation only, and not an interval from this test. This is intended to clarify the computer method for arriving at the VOC mass emitted data for each test interval.

Barometric Pressure (Baro-P) = 768.4 mm Hg
Exhaust Pressure (Exhaust-P) = 1.0 mm Hg
Ambient Temperature (Ambient-T) = 16.8° C
Exhaust Temperate (Exhaust-T) = 18.3° C
Volume Emitted Standardized (VES) = 43.6 m³
Volume

Volume Emitted (VE) = 42.9 m3
Milligrams Emitted (ME) = 436931.5 mg
Outlet VOC Concentration = 0.55 %
Inlet VOC Concentration = 34.0 %

<u>Please Note: All data fields are rounded to two places following the decimal point for display purposes only.</u>

1.) Therefore, for this calculation:

0.545 % lowest possible value before rounding for display

HCout = 0.55 % value displayed (after rounding)

0.554 % highest possible value before rounding for display

43.55 m³ lowest possible value before rounding for display

VES = 43.6 m³ value displayed (after rounding)

43.64 m³ highest possible value before rounding for display

2.) Using the above values in the previous equation B we have:

 $(1.83 \times 106) \times (0.00545) \times (43.55) = 434,345.9 \text{ mg}$ 436,931.5 mg

 $(1.83 \times 106) \times (0.0055) \times (43.6) = 438,834.0 \text{ mg}$ $(1.83 \times 106) \times (0.00554) \times (43.64) = 442,431.0 \text{ mg}$

Note: The value for ME printed by the computer for this interval is 436,931.5 mg. While this is not the result produced from entering the printed values for HCout and VES into Equation B, it is the result produced by the calculation carried out on the stored computer data, prior to rounding for display.

DATA SUMMARY

TERMINAL DESCRIPTION Chevron Products
Port Everglades, Florida

VAPOR CONTROL UNIT TYPE John Zink VRU

TEST DATE July 14, 2011

TEST PERIOD 07:30–13:30 Six Hrs

AVERAGE AMBIENT TEMPERATURE 32.84 Deg C

AVERAGE OUTLET CONCENTRATION (as Propane) 0.02 % by Volume

AVERAGE INLET CONCENTRATION (as Propane) 33.69 % by Volume

TOTAL PETROLEUM LOADED 267,308 gallons

ACCOUNTABLE PETROLEUM LOADED 216,100 gallons

AVERAGE HYDROCARBON EMISSIONS 0.23 mg/liter (Calculated with Total Loaded Product) 0.08 lb/hr

AVERAGE HYDROCARBON EMISSIONS

(Calculated with Accountable Product Loaded)

0.28 mg/liter

0.08 lb/hr

NUMBER OF TRUCKS LOADED 27 Gas Loads
Diesel Loads N/A

NUMBER OF LEAKING TRUCKS 0

VOLUME OF LEAKING TRUCKS 0 gallons


MAXIMUM PRESSURE AT TRUCK VAPOR HOSE 11.0" H₂0

STRIP CHART RECORDER SPEED 150 mm/hour


UNIT EFFICIENCY 99.95 %

COMPUTER PRINTOUT LEGEND


AMBIENT-TAmbient Ten	ıperature ^o Celsius
EXHAUST-T Exhaust Ten	aperatureº Celsius
EXHAUST-PExhaust Pres	suremm Hg
BARO-PBarometric I	ressuremm Hg
HCinInlet VOC 0	Concentration (when used)% by volume
HCoutExhaust VO	C ConcentrationVol. Fraction
VESFlow Through	h Turbine Meterm ³ std.
METotal Milligr	rams Emittedmg of VOC
VEFlow Through	gh Turbine Meterm ³



VAPOR RECOVERY UNIT TEST SCHEMATIC



Test Equipment Diagram Note: Not to Scale



APPENDIX A TRUCK MONITORING DATA SHEETS

Terminal Location	Cheuron POG+E	149/045	Date 7-14-2011	Page Number Z
Seq. No Bay No	Tanker Name PPEL Trailer Number 87	VINE THE EAST OFFICE PROPERTY AND ADDRESS OF THE PARTY.	Load Start Ti Load Stop Ti	me 8:37 AM me 8:50 AM
Products Loading	Previous Product	Leak	Accountable Gallons	Non-Acct. Gallons
GAS	GAS	NÕ	7000	0
Max. Back Pressur	e: Reading 1 Reading	g 2 <u> </u>	ading 3 5 Reading 4	L Highest)
Seq. No. 6 Bay No. 1 Products Loading	Tanker Name King (Trailer Number 803 (Previous Product) Advav Leak	(†σόμ) Load Start T Load Stop T <u>Accountable Gallons</u>	V - A - Ju L A - 1
GAS	GAS	\wedge	6000	0
Max. Back Pressu	e: Reading 1 Readin	g 2 Re	ading 3 3 Reading 4	니 Highest 니
Seq. No Bay No Products Loading	Tanker Name Mc 4-27 Trailer Number 4463 Previous Product	20	Load Start T	ime_8:54_AM ime_9:00AM Non-Acct, Gallons
GAS	GAS	40	8800	0
May Back Pressu	re: Reading 1_3 Readin	g 2 R	eading 3 6 Reading 4	6 Highest 8
Seq. No. 10 Bay No. 3 Products Loading	Tanker Name LEN OXY Trailer Number 50 L Previous Product	Havo	A HONOR Load Start T	ime 9:03 44 ime 9:13 44 Non-Acct, Gallons
GAS	GAS	N 0	8900	0
Harrison III and the second of	re: Reading 1 Reading	ng 2 DR	eading 3 Reading 4	Highest Sime 9.24AM
Seq. No. 11 Bay No. 2	Tanker Name_P.R. Liv Trailer Number_92	<u></u>	Load Stop T	Time 4.36AM
Products Loading	Previous Product	<u>Leak</u>	Accountable Gallons	Non-Acct. Gallons
GAS	GAS	NÕ	8700	0
Max. Back Pressu			eading 3 S Reading 4	
Seq. No. 17	Tanker Name_ABEN Trailer Number_1552	ADVI	Load Start Load Stop L	
Bay No Products Loading	Previous Product	Leak	Accountable Gallons	Non-Acct. Gallons
AVGAS	AVGAS	<i>M</i> 0	8000	0
Max. Back Pressu	re: Reading 1 Reading	ng 2 💆 R	eading 3_L Reading 4	B Highest 9
Accountable Accountable Accountable Accountable	Gallons Load 1 Gallons Load 2 Gallons Load 3 Gallons Load 4 Gallons Load 5 Gallons Load 6 Gallons Load 6		Total Gallons Load ?	3
Total Accountable Ga Acct. Total Fron	Previous Page + 17	500	Total Gallons This Pag Total Gallons Prev. Pag	
Accountabl	e Gallons Total $= 94/6$	100	Total Gallons	**************************************

Terminal Location_	Chevron Port	Everglode	Date 17-14-2011	magazita pyromani moznowa kontrologicki pod promozni promozni promozni za promozni promozna promozna promozna p	
Seq. No. 25 Bay No. 3	Tanker Name Pikeline Trailer Number 38		Load Start Ti Load Stop Ti	me 12:31 PM me 12:42 PM	
Products Loading	Previous Product	<u>Leak</u>	Accountable Gallons	Non-Acct, Gallons	
GAS	GA5		8700	0	
			eading 3 5 Reading 4	Hightest (place) of the second	
Seq. No. 26 Bay No. 2 Products Loading	Tanker Name 699 1 Trailer Number 232 1 Previous Product	Leak	Load Start To Load Stop To Accountable Gallons		
GAS	GAS	NO	8800	0	Section Control Section Contro
	77 17		Rendistrat	V) Highest D	
Max. Back Pressur Seq. No. 27	Tanker Name Kengu		eading 3 / Reading 4 hh 90 Load Start T	ime 1:14 PM	
Bay No	Trailer Number 504	[6	Load Stop T Accountable Gallons		
Products Loading	Previous Product	<u>Leak</u>	non particular in the control of the	1,4028 13032 (323023)	Commission of the Control of the Con
GAS	GAS		8 0 00		
Max. Back Pressu	1	ng 2R	eading 3 S Reading 4	The state of the s	
Seq. No Bay No	Tanker Name Trailer Number		Load Start T Load Stop T	ime	
Products Loading	Previous Product	<u>Leak</u>	Accountable Galions	Non-Acct. Gallons	
		The state of the s			
		Million Marian			
	re: Reading 1 Readi	ng 2 R	eading 3 Reading 4_ Load Start T	Highest	
Seq. No Bay No			Load Start 1 Load Stop T	ime	
Products Loading	***	<u>Leak</u>	Accountable Gallons	Non-Acct. Gallons	
ALCONOMINATION OF THE PROPERTY		STATE OF THE STATE		THE PROPERTY OF THE PROPERTY O	
				LES CONT	
Max. Back Pressu	A CONTRACTOR OF THE PARTY OF TH	ing 2 R	eading 3Reading 4_ Load Start T	Highest	
Seq. No Bay No	Tanker Name Trailer Number		Load Stop T	ime	CONTRACTOR OF THE PROPERTY OF
Products Loading	Previous Product	<u>Leak</u>	Accountable Gallons	Non-Acct. Gallons	d Statement
				OLA TOTAL PORTON	To the state of th
	CALCULATION OF THE PROPERTY OF				
Max. Back Pressu		<u> </u>	eading 3 Reading 4_	Highest	
	Gallons Load 1 8700 Gallons Load 2 880		Total Gallons Load 1 Total Gallons Load 2		DIESEL LOADE
	Gallons Load 2 <u>860</u> Gallons Load 3 <u>860</u>		Total Gallons Load 3		AT DIESEL PAG
	Gallons Load 4 Gallons Load 5		Total Gallons Load 4 Total Gallons Load 5		FROM CHENKON
	Gallons Load 6		Total Gallons Load 6		METERS.
Total Accountable Ga Acct. Total From	HORIS EIIIS EAGO	500	Total Gallons This Page Total Gallons Prev. Page	والمراجع والم والمراجع والمراجع والمراجع والمراجع والمراجع والمراجع والمراج	4

APPENDIX B

INSTRUMENT AND CALIBRATION INFORMATION

enviromaxTM

THE GOLD STANDARD IN GAS ANALYSIS

MODEL 2010

Nondispersive Infrared
Analyzer



Rack Mounted Single Gas Analyzer



Rack Mounted Dual Gas Analyzer

- Patented Detector Provides Outstanding Selectivity and Sensitivity
- Microprocessor Linearization Increases Accuracy
- All Digital Electronics Improves Performance
- Totally Closed Sample Path is Not Exposed to Ambient Conditions
- No Moving Parts to Adjust or Wear Out Improves Reliability
- Standard Automatic Calibration Assures Your Data is Accurate
- Two or Four Adjustable Analog Chart Recorder Ranges
- Adjustable High and Low Audible Alarms or Limits
- Selectable Voltage Output and Isolated 4-20 mA
- RS-232 and Digital Input/Output Options Available
- Slide-out Drawer Provides Complete Access to All Internal Components
- Optional Oxygen Channel Provides Two or Three Gas Capability in One Analyzer



The Model 2010 optical bench consists of the IR source, sample cell, and patented detector. The light source is electronically chopped and replaces the need for a mechanical chopper thereby eliminating all moving parts from the optical bench. The totally closed sample path eliminates interfering absorption from ambient air and the need for periodic purging of the instrument case. The sample cells incorporate a unique design in size and shape allowing the path length of the sample cell to be optimized to the measuring range of the gas of interest, reducing noise and increasing sensitivity. The patented Model 2010 detector consists of two chambers in optical series with a sensitive flow transducer to measure the relative infrared energy absorbed by each chamber. The primary difference between the Model 2010 detector and other NDIR detectors is that the signal between the primary and secondary chambers is balanced. This balancing process greatly improves the sensitivity and selectivity for the gas of interest. Other improvements include very low drift rate and minimal response to vibration and temperature changes.

The Model 2010 is extremely simple and easy to operate. All operating parameters and conditions are controlled from four front panel buttons. The Operator can control such settings as calibration gas concentration, calibration frequency, chart recorder range, flush time, and alarm conditions. All settings are protected by battery backup in case of loss of power. The Model 2010 also has an extensive diagnostic program accessed from the front panel for troubleshooting purposes.

Chopped
Light Source

Sample
Cell

Patented
Balanced

Model 2010 Single Gas Analyzer shown with slide-out drawer open

Standard Gases and Typical Ranges:

Gas	Low Range From:	High Range From:
CO	0-50 PPM	0-100%
CO_2	0-10 PPM	0-100%
SO_2	0-5000 PPM	0-100%
NO	0-500 PPM	0-100%
N20	0-10 PPM	0-100%
CH ₄	0-50 PPM	0-100%
C_3H_8	0-50 PPM	0-100%
C_2F_6	0-50 PPM	0-100%

Other gases available upon request.

Typical Performance*

Low ppm Rang	es Other Ranges	
Resolution	0.1 PPM	0.1%
Repeatability	0.5% or better	0.1% or better
Noise	0.5% or better	0.1% or better
Drift/24 hrs.	1.0% or better	0.3% or better

^{*}Expressed as % of full scale unless otherwise indicated. At constant ambient conditions and with daily zero calibration. All specifications are subject to change without notice.



Detector

18900 Teller Avenue Irvine, CA 92612-1617 Phones: (949) 756-1632

(800) 927-4572

Web: www.listonscientific.com

Fax: (949) 756-1635 e-mail: info@listonscientific.com

MODEL 2010 Nondispersive Gas Analyzer

General Specifications*

Measuring method: NDIR single beam

Sample flow: Standard, 0.2-2.0 liters/minute

Low ranges, 2.0-10.0 liters/minute

Gas measured: User specified

Sample temperature: -10 to +50° C

Measuring range: See chart on individual product sheets

Dust: Should be dust-free

Display: Vacuum Fluorescent, 2 lines of 16 characters

Moisture: Non-condensing

Alarms: High & Low limit, user selectable

Pressure: Inlet pressure not to exceed 5 PSIG

Analog output: Selectable 1,5 or 10 volts

Exhaust gas: Non-restrictive outlet aperture and hoses

Analog ranges: 4, each with selectable full scale and auto-ranging

Output options: RS-232, Digital Input/Output

Warm-up time: Usable in 1 hour. Optimum in 3 hours

Front panel: 4-button operator interface

Operating temperature: -10 to 50° C

Cal valve actuation: Isolated triac control, maximum load .6 amp

Power source: 120/230 VAC., 50/60 Hz. user specified.

Portable: battery charger, minimum 6 hours operation

Materials in sample path: Glass, Gold, Buna-N, Lexan, Epoxy,

Sapphire, Zinc Sulfide or Calcium Fluoride, 304 SS

Power consumption: 50 watts per channel

 * All specifications are subject to change depending upon gas and range; and may be change without prior notice.

Dimensions and Weight

Rack Mounted Version

Length: 22.5 in (57.1 cm)

Width: 17.1 in (43.4 cm)

Height: 5.25 in (13.3 cm)

Weight: 38 lbs (17.3 kg)

Portable Version

Length: 20.0 in (50.8 cm)

Width: 8.5 in (21.6 cm)

Height: 5.25 in (13.3 cm)

Weight: 19 lbs (8.6 kg)

Explosion Proof Version

Electronics Enclosure

Length: 15 in (38.1 cm) Width: 7.5 in (19.05 cm) Height: 7.5 in (19.05 cm) Weight: 13 lbs (5.9 kg)

Explosion Proof Enclosure

Length: 10 in (25.4 cm) Width: 15 in (38.1 cm)

Height: variable, 15.5 in (39.4 cm)

to 27.5 in (69.9 cm)

Weight: variable, 51 lbs (23.2 kg), 63 lbs (28.6 kg) or 75 lbs (34.1kg)



18900 Teller Avenue Irvine, CA 92612 949-756-1632, 800-927-4572

Fax: 949-756-1635 email: info@listonscientific.com Web: www.listonscientific.com

HORIBA ENVIRONMENTAL AND PROCESS INSTRUMENTS

News Company Service **Financing** Contact **Products**

VIA-510 Gas Analyzer

Features

- Selectable response time
- Selectable outputs: 0–1 VDC or 4–20 mA
- Digital outputs indicate malfunctions or calibration failure)
- Measures CO, CO2, NOx, SO2, CH4, C2H4, and NO2; others upon request



Overview

The VIA-510 series of general-purpose gas analyzers provide continuous monitoring of concentrations of the specific sample gas. The analyzers can be operated from controls on the front panel or by commands from a remote computer. Measurement results are displayed on the front panel and are available to remote data logging systems through an industry-standard interface.

The VIA-510 series can be used for a wide variety of analyses and tests, such as industrial process control and composition analysis, environment-related atmospheric and fixed-source emissions monitoring, and automobile emission analysis.

These analyzers use the infrared absorption method which offers superior sensitivity, selectivity, and stability.

They are compact and compatible with a variety of OEM analysis equipment.

A high level of sensitivity is achieved through the use of a dual-beam NDIR analysis method.

Horiba's patented

chopper motor assures continuous long-term stable monitoring. The analysis mechanism and the amplifier are combined in a single unit. The highly accurate performance makes the analyzers suitable for process monitoring and control.

Specifications

Standard Ranges

Maximum Minimum Gas 0-100% Carbon monoxide 0-50 ppm

(CO)		
Carbon dioxide (CO2)	0-50 ppm	0-100%
Nitrogen monoxide (NO)	0-100 ppm	0-100%
Sulfer dioxide (SO2)	0-100 ppm	0-100%
Methane (CH4)	0-100 ppm	0-100%
Ethene (C2H4)	0-100 ppm	0-100%
Nitrous Oxide (N2O)	0-100 ppm	0-100%

Performance

Lowest detection limit: 1.0 ppm

Repeatability:

± 1% of full-scale

Response time:

Selectable

Zero drift:


< 1% (full scale) per day

Span drift:

< 2% (full scale) per week

Copyright 2003 Horiba Inc.

AMERICAN METER GTX Gas Turbine Meters Data Sheet



American Meter's gas turbine meters bring together refinements in design learned from years of application experience. GTX turbine meters are based on the standard GTS turbine meter design with the removal of certain features, making them an extremely cost-effective option for industrial applications while maintaining meter accuracy and the quality expected from American Meter products.

GTX meters are available in 4", 6", and 8" sizes with maximum allowable operating pressures up to 175 psig. As shown in the features and benefits listed, many of the optional accessories for GTS turbine meters remain available on the GTX Series, such as medium- and low-frequency pulsers.

Related Bulletins:

Instruction Manual				R						**			.IM 4720
Repair Parts List								je	*	*		•	.RPL 4810
Medium Frequency		Į	ı	5	e	r	4			=			JMP 6778
Electronic Temp. Co	ď	n	p		_				,	•			.IM 4730

Features - Benefits - Optional Accessories

- Mechanical Drive Output
- Mechanical and Electronic Pulse Outputs
- Temperature Compensation
- Mechanical Drive Models for use with P&T Correctors or Electronic Pulse Output Models to interface directly with popular flow computers.
- Electronic Temperature Compensation with Fixed Factor Pressure and six-month data storage.
- One Output Gear Train for all meter sizes; reduces spare parts inventory.
- High-Efficiency Inlet Flo-Guide⁶ flow conditioners to minimize the effects of flow disturbances in shortcoupled installations.
- Interchangeable Pre-Calibrated Measurement Cartridges for easy maintenance.
 NOTE: GTX cartridges fit only into GTX bodies.
- · Three-Point Accuracy Curve supplied as standard.
- · Five-Point Accuracy Curve (optional).
- Medium- and High-Pressure Accuracy Curves available.
- · Cartridge Recertification and Repair Services.
- · Mercury or Equimeter Corrector Adapter Plates.
- Output Drive: 100 ft (4" and 6")
 Output Drive: 1,000 ft (8")

Comparison Chart

Feature	GTS	GTX
Pressure Ratings	175, 720, 1,440	175
Bearings/	Standard SST Bearings/Oller	a
Outlet Diffuser	Standard	None
Compatibility	GTS/AccuTest/GT	GTX Only
Rotors	45° or 30° Metal or Plastic	
Pulsers	High-, Medium-, or Low-Frequency	

Capacity Table

			45° Rotor Angle			
Size	Qmax Qmin MSCFH MSCFH		Range Qmax/Qmin	Minimum Actual Flow Rate MACFH	Pressure Drop Inches W.C.	
411	18	1.2	15	1,20	2,4	
6 ¹⁰	35	1.9	18	1.94	3.3	
8"	60	3	20	3.00	1.6	

GTX Basic Specifications (Figure 1)

				1		Flan	ge				
	1						Во	lts			
Size	Mate AL=alu		Dimer (incl	CONTRACTOR STATE	(Inches)		nches) Dia.			Weight	Cartridge Bolt Torque
	Body	Тор	A	В	O.D.	*B.D.	No.	(in.)	ANSI	(lbs.)	(lb-ft.)
40	AL	AL	5.85	14.0	9.00	7.50	8	5/8	150 FF	32	20
6н	AL	AL	6.42	16.0	11.00	9.50	8	3/4	150 FF	54	35
8"	AL	AL	7.42	21.0	13.50	11.75	8	3/4	150 RF	90	70

Note: GTX MAOP = 175 psig *B.D. = Bolt Circle Diameter

B = Standard **Operating Temperature Range:**

-40°F to + 140°F -40°C to -60°C

Manufacturing Standards ANSI/ASME MFC - 4M - 1986 AGA Report #7




Figure 1 **GTX Specifications**

GTX Ordering Information

Maximum Allowable Working Pressure: Size:

175 psi/12 bar 4" - 100 mm 175 psi/12 bar 6" - 150 mm 175 psi/12 bar 8" - 200 mm

Index: Clock Type, Odometer Type, None

Model:

Mechanical Instrument Drive Output only Mechanical Instrument Drive Output and Medium-Frequency Pulse Output

- Mechanical Instrument Drive and **Low-Frequency Pulse Output**
- Mercury or Equimeter Corrector **Adapter Plate**



AMERICAN METER

300 Welsh Road **Building One** Horsham, PA 19044-2234 U.S.A. Phone: 215/830-1800

Fax: 215/830-1890

Website: americanmeter.com



CANADIAN METER

275 Industrial Road Cambridge, Ontario, N3H 4R7 Canada Phone: 877/461-2626 (toll free)

Phone: 519/650-1900 Fax: 519/650-1917

Website: canadianmeter.com





ISO 9002 Registered



International Measurement and Control Systems

90 Main Street • P.O. Box 1605 Tullytown, Pennsylvania 19007 Tel: 215-946-2200 • Fax: 215-943-2984


E-mail: sales@imacsystems.com

Web site: www.imacsystems.com

GAS METER INSPECTION CERTIFICATE

2/24/2011 Date: T - 60Meter: 1000 cu.ft. Drive Rate: 813171 Serial Number: LD. Type of Proof: 51717 Work Order#: 68 Meter Temp [F]: 175 Leak Test [psi]: N/A As Left - Unc: 175 P Max [psi]: N/A As Left - Cor: Customer Name: Jordan Technologies

Percent of Max. Flow	Flow [ft3/h]	Corrected Proof	Diff. Press. [inW.C.]	Uncorrected Proof
20.0	12000	100.4	1.3	
20.0	12000	100.4	1.3	
15.0	9000	100.3	0.7	
15.0	9000	100.3	0.7	



Attested to by

3/15/2010







D/Advanced.

DXAdvanced DX1000N Removable Chassis Model

A removable chassis model has been added to the Yokogawa's latest DXAdvanced Video Graphic Recorder featuring easy maintenance.

The new DXAdvanced DX1000N features an inner chassis that can be removed from the case via the front panel of the instrument. This provides access to all of the internal components of the DX1000N from the control panel without having to access the rear of the unit or disturb any of the field and power supply wiring. Functionality, appearance, and panel cutout dimensions are the same as those of the standard DX1000.

Advanced Performance

- High-speed measurement
- * High-speed measurement of up to 25 ms (DX1002N or DX1004N using fast sampling mode)

Advanced Memory

- High Capacity Internal Memory and Removable Media
- * Supports up to 200 MB of non-volatile, internal flash memory for reliable, long-term data storage
- * All models include a CompactFlash drive. Rugged and readily available CompactFlash cards (CF cards) serve as the removable media, and are available as optional accessories.
- * Supports USB Flash drive with optional USB interface.

Advanced Display and User Interface Functions

- Easy configuration and menu navigation
- * USB keyboard & remote control options for text entry
- * Versatile, standard display modes
- * Jump to your favorite screen with the Favorite key

Advanced Connectivity

- Powerful Ethernet connectivity and convenience functions
- * Standard Ethernet interface
- Includes Web server and E-mail messaging functions, time synchronization (SNTP), automatic network setup (DHCP), file transfer (FTP) and more.

Bulletin 04L43B01-01E

Advanced Reliability and Security

- Rugged construction and data security
- * Water and dust-proof front panel (complies with IEC529-IP65 and NEMA No.250 TYPE4*) *Except for external icing test.
- * A mechanical lock with removable key is provided to securely latch the front panel door. This forbids access to the power switch and removable media.
- * Reliable, non-volatile flash memory is used for internal data storage operations with ECC* function. * ECC: Error Check and Correct

www.dagetation.com















Application Software (DAQSTANDARD for DXAdvanced)

Every DXAdvanced unit includes a DAQSTANDARD software, which is used for all data file display and reporting functions, including printing and conversion to common file formats. In addition, it includes a configuration tool that is used to fully configure the unit in both on-line (via Ethernet communications) and off-line (saving and loading files from the media) modes. Configuration files can also be archived on the PC.



Models and Suffix Codes

	Suf		Optiona code	
DX1002N				2 ch, 125 ms (Fast sampling mode; 25 ms)
DX1004N				4 ch, 125 ms (Fast sampling mode: 25 ms)
DX1006N				6 ch, 1 s (Fast sampling mode: 125 ms)
DX1012N				12 ch, 1 s (Fast sampling mode: 125ms)
Internal	-1			Standard memory (80 MB)
memory	-2			Large memory (200 MB)
External media	4			CF card (with media)
Display langua	ige	-2		English/German/French, degF, DST (summer/winter time)
Options		18	11	Alarm output 2 points *1
		10	12	Alarm output 4 points *1
		18	13	Alarm output 6 points *1 *2
/C2 /C3 /F1		2	RS-232 Interface *3	
		23	RS-422A/485 interface *3	
		1	FAIL/Status output *2	
		/1-	12	Clamped input terminal (detachable)
		//\	/11	Mathematical functions
		/	11	Cu10, Cu25 RTD input/3 leg isolated RTD
		//	12	3 leg isolated RTD *4
		1000	13	Extended input type (PR40-20, Pt50, etc.)
		/F	1	24 VDC/AC power supply
		15	31	Remote control
		-	PS2	24VDC transmitter power supply (2 loops) "5
		Л	PS4	24VDC transmitter power supply (4 loops) *6
/KB1		(B1	Easy text entry (with input terminal) *7 *8	
		100	(B2	Easy text entry (without input terminal) *7
		10.7	JSB1	USB interface
		/F	PM1	Pulse input (including remote control and mathematical functions) *9
		10	C1	Calibration correction function

"1/A1, /A2 and /A3 cernot be specified together. "2 /A3 and /F1 cannot be specified together.
"3 /C2 and /C3 carnot be specified together. "2 /A3 and /F1 cannot be specified together.
"5 /C2 and /C3 carnot be specified together. "8 /A2 /A3 or /F1 carnot be specified together."
"5 in case that /TP84 is specified, /TP82, /A, /A2 /A3 or /F1 carnot be specified together."
"7 /K61 and /K62 carnot be specified together." Is no case that /R61 is specified, renote input terminal (438227) is included.
"9 in case that /FA1 is specified, /A3, /A11, /R11, /TP82 or /TP84 carnot be specified. And combination of /A2/F1 cannot be specified together."


Application Software

Description	O S
DAQSTANDARD for DXAdvanced	Windows 2000/XP

Accessories

	Model code (part number)	Specification
	415920	25011±0.1%
Shunt resister	415921	100Ω±0.1%
(for screw input terminal)	415922	1011±0.1%
	438920	25011±0.1%
Shunt resister (for clamped	438921	100 (1±0.1%
input terminal)	438922	10Ω±0.1%
CF card adapter	772090	
	772091	128 MB
	772092	256 MB
CF card	772093	512 MB
	772094	1 GB
Mounting bracket	B9900BX	-
Door lock key	B8706FX	
Remote control terminal	438227	For /KB1, /KB2 option

Dimensions



For more details on all functions, see the DX1000/DX2000 catalog (Bulletin 04L41801-01E). For more details on all specifications, see the DX1000N General Specifications (GS 04L43801-01E).

Daqstation and DXAdvanced are registered trademark of Yokogawa Ele Microsoft, MS and Windows are registered trademarks or trademark metusur, set and vinnous are registered trademarks or trademarks of Milcro States and other countries. Ethernet is a registered trademark of Xerox Corporation. Other company names and product names appearing in this document are registed of their respective holders.









What does Yokogawa vigilance mean to the future of your business? Quality: Through products that are built from the ground up and tested to the last hour, you're ensured continuous operation and more uptime. Innovation. Your business will benefit from new insights and copobilities, bringing two predictability to your process. Foresight. As the market changes, you'll have solutions that give you the continuity and flexibility to plan chead and grow. Our partners know the difference. With Yokogawa, you can caunt on a lifetime of plant efficiency, from instrumentation to operation support. Let us be vigilant about your business.

YOKOGAWA ELECTRIC CORPORATION

Network Solutions Business Div./Phone: (81)-422-52-7179, Fax: (81)-422-52-6619

E-mail: ns@cs.jp.yokogawa.com **YOKOGAWA CORPORATION OF AMERICA**

YOKOGAWA EUROPE B.V.

YOKOGAWA ENGINEERING ASIA PTE. LTD.

Phone: 800-888-6400, Fax: (1)-770-251-6427 Phone: (31)-33-4641806, Fax: (31)-33-4641807

Phone: (65)-62419933, Fax: (65)-62412606

NetSOL Omline Sign up for our free e-mail newsletter

Vig-RS-1E Printed in Japan, 702(KP) [Ed:03/b]





GUARANTEED 3 DAY SHIPMENT FOR STANDARD PARTS (loss than 10 piscas)

Model 264

Very Low Differential Pressure Transducer

Unidirectional Ranges: 0 - 0.1 to 0 - 100 in. W.C. Bidirectional Ranges: 0 - ± 0.5 to 0 - ± 50 in. W.C.

Air or Non-Conducting Gas



sense differential or gauge (static) pressure and convert this pressure difference to a proportional electrical output for either unidirectional or bidirectional pressure ranges. The 264 Series is offered with a high level analog 0 to 5 VDC or 4 to 20 mA output.

Used in Building Energy Management Systems, these transducers are capable of measuring pressures and flows with the accuracy necessary for proper building pressurization and air flow control.

The 264 Series transducers are available for air pressure ranges as low as 0.1 in. W.C. full scale to 100 in. W.C. full scale. Static standard accuracy is ±1.0% full scale in normal ambient temperature environments, but higher accuracies are available. The units are temperature compensated to 0.033% FS/°F thermal error over the temperature range of 0°F to +150°F.

The Model 264 utilizes an improved all stainless steel micro-tig welded sensor. The tensioned stainless steel diaphragm and insulated stainless steel electrode, positioned close to the diaphragm, form a variable capacitor. Positive pressure moves the diaphragm toward the electrode, increasing the capacitance. A decrease in pressure moves the diaphragm away from the electrode, decreasing the capacitance. The change in capacitance is detected and converted to a linear DC electrical signal by Setra's unique electronic circuit.

The tensioned sensor allows up to 10 PSI overpressure (in either direction) with no damage to the unit. In addition, the parts that make up the sensor have thermally matched coefficients, which promote improved temperature performance and excellent long term stability.

NOTE: Setra quality standards are based on ANSI-Z540-1. The calibration of this product is NIST traceable.

U.S. Patent nos. 4093915; 4358814; 4434203; 6019002; 6014800. Other Patents Pending.

Applications

- Heating, Ventilating and Air Conditioning (HVAC)
- Energy Management Systems
- Variable Air Volume and Fan Control (VAV)
- Environmental Pollution Control
- Lab and Fume Hood Control
- Oven Pressurization and Furnace Draft Controls

Features

- Up to 10 PSI Overpressure on All Ranges
- Installation Time
 Minimized with Snap
 Track Mounting and
 Easy- To-Access Pressure
 Ports and Electrical
 Connections
- 0 to 5 VDC or 2-wire 4 to 20 mA Analog Outputs Are Compatible with Energy Management Systems
- Reverse Wiring Protection
- Internal Regulation
 Permits Use with
 Unregulated DC Power
 Supplies
- Meets (Conformance Standards

When it comes to a product to rely on-choose the Model 264. When it comes to a company to trust-choose Setra.



Visit Setra Online: http://www.setra.com



Performance Data

	Standard	Opti	onal
Accuracy* RSS(at constant temp)	±1.0% FS	±0.4% FS	±0.25% FS
Non-Linearity, BFSL	±0.96% FS	±0.38% FS	±0.22% FS
Hysteresis	0.10% FS	0.10% FS	0.10% FS
Non-Repeatability	0.05% FS	0.05% FS	0.05%FS
Thermal Effects**			
Compensated Range F(°C)	0 to +1	50 (-18 to	+65)
Zero/Span Shift %FS/°F(°C)	0.033 (0.06)	
Maximum Line Pressure	10 psi		
Overpressure	Up to 10 psi in Positive or		
5.55.	Negativ	e Direction	
Long Term Stability	0.5% F	S/1 YR	
		Zei	ro Offset
Position Effect	Range	(9	%FS/G)
(Unit is factory calibrated at ()q To 0.5	in. WC	0.60
effect in the vertical position	.) To 1.0	in.WC	0.50
	To 2.5	in.WC	0.22
	To 5 ir	. WC	0.14

^{*} RSS of Non-Linearity, Hysteresis, and Non-Repeatability.

Model 264 Specifications

Environmental Data

Temperature

Operating F (°C)

0 to +175 (-18 to +79)

-65 to +250 (-54 to +121) Storage F (°C)

Operating temperature limits of the electronics only. Pressure media temperatures may be considerably higher.

Physical Description

Fire-Retardant Glass Filled Case

Polyester

Four screw holes on removable Mounting zinc plated steel base (designed

for 2.75" snap track)

Electrical Connection Pressure Fittings

Screw Terminal Strip 3/16" O.D. barbed brass pressure fitting for 1/4" push-on

tubing

Zero and Span Adjustments Accessible on top of case 10 ounces

Weight (approx.)

Pressure Media

Typically air or similar non-conducting gases.

Specifications subject to change without notice.

Electrical Data (Voltage)

3-Wire (Com, Exc, Out) Circuit

Excitation 9 to 30 VDC O to 5 VDC Output*

Bidirectional output at zero

2.5 VDC" pressure: 100 ohms Output Impedance

*Calibrated into a 50K ohm load, operable into a 5000 ohm load or greater. *Zero output factory set to within ±50mV (±25 mV for optional accuracies).
*Span (Full Scale) output factory set to within ±50mV. (±25 mV for optional accuracies).

Electrical Data (Current)

2-Wire Circuit 4 to 20mA" Output

Bidirectional output at zero

12mA" pressure:

0 to 800 ohms External Load Minimum supply voltage (VDC) = 9+ 0.02 x

(Resistance of receiver plus line).


Maximum supply voltage (VDC) = 30 + 0.004 x


(Resistance of receiver plus line).

"Calibrated at factory with a 24 VDC loop supply voltage and a 250 ohm load.
"Zero output factory set to within ±0.16mA (±0.08 mA for optional accuracies).

**Span (Full Scale) output factory set to wtihin ±0.16mA (±0.08 mA for optional accuracies)


Outline Drawings





Code all blocks in table.

Example: Part No. 26412R5WD11T1C for a 264 Transducer 0 to 2.5 in. WC Range, 4 to 20 mA Output, Terminal Strip Electrical Connection, and ±1% Accuracy.





^{**}Units calibrated at nominal 70°F. Maximum thermal error computed from this datum.



Vacuum Differential

PSID Ranges

DATUM 2000™

Manometer w/Transducer Installed

The DATUM 2000™ Manometer is a complete system with built-in pressure or vacuum transducer. A wide selection of standard pressure ranges is available for gauge, absolute, differential and vacuum measurements from ±0.25 inches WC to 1000 psi. The transducer installed in the Manometer operates on channel one of the two channel meter. Channel two is available for any other voltage or current input from another remote instrument. Setra pressure transducer Models 204, 204D, 239 and 270 are available for installation in the 1/4 DIN Datum Manometer.

Barometric

DATUM	2000™ Manometer	Transducer	Specifications
	w/Models 204/204D	w/Model 239	w/Model 270
Type of Pressure Measurement	Gauge	Differential	Gauge
ivieasur ement	Absolute	Gauge	Absolute

Standard Ranges

0 to 25, 50, 100, 250,	0 to 0.5, 1.0, 2.5, 5,	0 to 5, 10, 20, 50,
500,1000 psig	15, 30 inch WC	100 psig
0 to 25, 50, 100, 250,	0 to ± 0.25 , $\pm .5$,	0 to 10, 20, 50,
500, 1000 psia	±1.0, ±2.5, ±7.5,	100 psia
0 to 25, 50, 100 psid	±15 inch WC	600-1100 millibar
0 to 14.7 psiv	0 to 5, 10 psid	800-1100 millibar
0 to ±10, ±25, ±50,	0 to ±2.5, ±5 psid	
±100 psid		

System Accuracy (RSS)

±0.11% FS ± 2 digits	±0.14% FS ± 2 digits	±0.05% FS ± 2 digits
±0.22% FS ± 2 digits*		
* For +100, +250, +500		

Thermal Effects

%FS +60° to +95°F			
Thermal Zero Shift	0.14 max. ± 4 digits	0.35 max. ± 4 digits	$0.07 \text{max.} \pm 4 \text{digits}$
Thermal Span Shift	0.11 max. ± 4 digits	$0.35 \text{max.} \pm 4 \text{digits}$	0.04 max. ± 4 digits
Pressure Fittings			

3			AND THE RESIDENCE OF THE PROPERTY OF THE PROPE
Positive	1/4"-18 NPT internal	1/8"-27 NPT internal	1/8"-27 NPT interna
Reference	1/8"-27 NPT internal	1/8"-27 NPT internal	N/A

Pressure Media

Positive	Gas compatible	Gases compatible	Non-condensing air
	with 17-4 PH	with stainless steel,	or gas compatible
	stainless steel.**	hard anodized 6061	with aluminum,
	**Note: Hydrogen not	aluminum,	alumina, ceramics,
	recommended for	Buna N O-ring.	gold, fluorocarbon
	use with 17-4 PH		elastomer sealant
	stainless steel.		and Buna-N O-Ring.

Reference

Clean dry air or other Clean dry air or other gases. (Non-corrosive, gases. (Non-corrosive, non-condensable.) non-condensable.)

N/A


Analog Output


Normally 0 to 5 VDC for unidirectional pressure or vacuum ranges. 0 to ±2.5 VDC for bidirectional ranges.

Normally 0 to 5 VDC for unidirectional pressure. 0 to ±2.5 VDC for bidirectional ranges.

0 to 5 VDC for gauge and absolute ranges.







Ordering Instructions

DATUM 2000™ Meter only

Order as DATUM 2000-1 meter for 115 VAC converter or DATUM 2000-2 for 220 VAC converter with European 2-prong turret.

DATUM 2001 Meter with One Transducer or Transmitter Set-up

To order factory set-up with one transducer or transmitter and 10ft. cable/connector assembly, specify option 2001-1 for 115 VAC converter or 2001-2 for 220 VAC European converter.

Transducer or transmitter ordered and priced separately.

DATUM 2002 Meter with Two Transducers or Transmitters Set-up

For two factory setups and cable assemblies with two transducers or transmitters, specify option 2002-1 for 115 VAC or 2002-2 for 220 VAC European converter.

Transducers or transmitters ordered and priced separately.

DATUM Manometer with Transducer Installed

To order a manometer with a Model 204 pressure transducer, order part #2204; with a Model 239, order part #2239; with a Model 270, order part #2270. Specify pressure range.

Options:

602: 1-5 VDC Output

(2204, 2239 only)

603: 1-6 VDC Output

(2204, 2239 only)

607: 0-5 VDC Output

(2239 bidirectional only)

653: 220 VAC converter (Manometer only)

654: RS-232 Output

811-825: 11-25 ft. of cable*

*Consult factory for lengths above 25 ft. of cable.

Specifications subject to change without notice.



159 Swanson Road Boxborough, Massachusetts 01719 Telephone: 978-263-1400 • Fax: 978-264-0292


800-257-3872



ONE TO SIX GAS PORTABLE MONITOR

Gas Detection For Life

EAGLE™ Model



Features

- Simultaneous detection of up to 6 different gases
- Over 250 gas monitoring configurations
- Wide range of toxic gases
- PPM / LEL hydrocarbon detection
- Powerful long-life pump up to 125' range with filters
- Low flow pump shut off and alarm
- Methane elimination switch for environmental use
- Security "Adjustment Lockout Switch"
- . Up to 30 hours of continuous operation
- Alkaline or Ni-Cad capability
- IR Sensors available for 50% CO₂,
 100% LEL CH₄, and 100% volume CH₄
- Transformer testing version available
- Datalogging option
- Autocalibration
- Dual hydrophobic filters (most versions)
- Ergonomic RFI / EMI / chemical / weather resistant enclosure
- Intrinsically safe design, CSA (C / US) & UL Classified (most versions)

RKI is proud to offer the most versatile portable gas detector on the market. Equipped with features that are not available on most competitive units, the EAGLE is a powerful instrument that does more than offer standard confined space protection. Detection combinations never before offered in a portable gas monitor are now available featuring the industry's widest selection of high quality, long life and field proven sensors.

The EAGLE features include PPM or LEL hydrocarbon detection at the push of a button, infrared sensors for CO2 and combustible monitoring including 100%volume methane, a methane elimination switch for environmental applications, a long list of super toxic gases and measurable ranges, and dual hydrophobic filters to increase its water resistant performance. For quick response and recovery from distant sampling locations, the EAGLE has a strong internal pump with a low flow auto shut off and alarm, which can draw samples up to 125 feet even with the dual hydrophobic filters in place. The EAGLE will continuously operate for over 30 hours on alkaline batteries or 18 hours on Ni-Cads. A variety of accessories are also available to help satisfy almost any application such as long sample hoses, special float probes for tank testing, datalogging, continuous operation adapters, remote alarms and strobes, and dilution fittings just to name a few.

With its ergonomic design and large glove friendly buttons, the EAGLE offers easy access to controls such as autocalibration, alarm silence, demand zero, peak hold and a wide variety of other features. Each channel has two alarm levels plus TWA and STEL alarms for toxic channels. The two alarm levels are user adjustable and can be latching or self resetting. Rugged, reliable, easy to operate and maintain, the EAGLE is the solution for just about any portable gas monitoring situation.

RKI Instruments, Inc · 33248 Central Ave. Union City, CA · Phone (800) 754-5165 · (510) 441-5656 · Fax (510) 441-5650

EAGLE™ Model

Enclosure	Weather resistant, chemical resistant, RFI / EMI coated high impact polycarbonate-polyester blend. Can operate in rain or set into 2.5" of water without damage. Ergonomically balanced with rugged top mounted handle.
Dimensions	10.5" L x 5.9" W x 7" H
Weight	5 lbs
Detection Principle	Catalytic combustion, electrochemical cell, galvanic cell, and infrared.
Sensor Life	2 years under normal conditions.
Sampling Method	Powerful, long-life pump (over 6,000 hours) can draw samples over 125 feet. Flow rate approximately 2.0 SCFH.
Display	4×20 LCD readout. Viewed through window in case top. Displays readings & status of all channels simultaneously. Backlight, automatic for alarms and by demand with adjustable time.
Alarms	2 alarms per channel plus TWA and STEL alarms for toxics. The two alarms are fully adjustable for levels, latching or self reset and silenceable.
Alarm Method	Buzzer 85 dB at 30 cm, dual high intensity LEDs, and flashing display.
Controls	6 external glove friendly push buttons for operation, demand zero, and autocalibration. Buttons also access LEL / ppm, alarm silence, peak hold, TWA / STEL values, battery status and many other features.
Continuous Operation	30 hours minimum using alkaline batteries, or 18 hours using Ni-Cads.
Power Source	4 alkaline or Ni-Cad, size D batteries. Charger has alkaline recognition to prevent battery damage if charging is attempted with alkalines.
Operating Temp. & Humidity	-10°C to 40°C (14°F to 104°F), 0 to 95% RH, non-condensing.
Indication Accuracy	Maximum variance +/- 5% of full scale.
Response Time	30 seconds to 90% (for most gases) using standard 5 ft hose.
Safety Rating	Intrinsically Safe, Class I, Division 1, Groups A, B, C and D. CSA (C / US) & UL Classified (most versions).
Standard Accessories	Shoulder strap, alkaline batteries, hydrophobic probe and 5 foot hose Internal hydrophobic filter (most versions) (certain toxic versions equipped with special probe, inlet fitting and 3' teflon hose. For HF and O3 versions, 3' teflon hose used without probe).
Optional Accessories	 Datalogging of up to 4 gases (No datalogging possible on 5 or 6 gas versions or versions with more than 2 toxic sensors) Remote alarms Dilution fitting (50/50) Ni-Cad batteries Battery charger, 115 VAC, 220 VAC, or 12 VDC Continous operation adapter, 115 VAC or 12 VDC Extra loud buzzer Extension probes Large internal hydrophobic filter
Warranty	One year material and workmanship

Specifications subject to change without notice.

Gases & Detectab	le Ranges					
Standard Confined Space Gases						
Hydrocarbons (CH ₄ , std)	0 - 100% LEL 0 - 50,000 ppm					
Oxygen (O ₂)	0 - 40% Vol.					
Carbon Monoxide (CO)	0 - 500 ppm					
Hydrogen Sulfide (H ₂ S)	0 - 100 ppm					
Super Toxics and Of	ther Gases					
Ammonia (NH ₃)	0 - 75 ppm					
Arsine (AsH ₃)	0 - 1 ppm 0 - 200 ppb					
Carbon Dioxide (CO ₂) (I R Sensor)	0 - 5,000 ppm 0 - 10,000 ppm 0 - 5% Vol. 0 - 20% Vol. 0 - 60% Vol.					
Chlorine (Cl ₂)	0 - 3 ppm					
Chlorine Dioxide (CIO ₂)	0 - 1 ppm					
Fluorine (F ₂)	0 - 5 ppm					
Hydrogen Fluoride (HF)	0 - 9 ppm					
Hydrogen Chloride (HCI)	0 - 15 ppm					
Hydrogen Cyanide (HCN)	0 - 30 ppm					
Hydrogen Sulfide (H ₂ S)	0 - 1 ppm 0 - 30 ppm					
Methane (CH ₄) (IR Sensor)	0 - 100% LEL 0 - 100% Vol.					
Isobutane (C ₄ H ₁₀) (IR Sensor)	0 - 100% LEL 0 - 30% Vol.					
Nitrogen Dioxide (NO ₂)	0 - 15 ppm					
Nitric Oxide (NO)	0 - 100 ppm					
Ozone (O ₃)	0 - 1 ppm					
Phosphine (PH ₃)	0 - 1 ppm					
Silane (SiH ₄)	0 - 15 ppm					
Sulfur Dioxide (SO ₂)	0 - 10 ppm 0 - 15 ppm					

The EAGLE can be configured with up to 6 gas sensors including a maximum of 2 super toxics from the above list.

Special Features

- Low flow alarm shuts pump off to avoid damage to instrument.
- Hydrophobic filter disc in probe.
- Internal hydrophobic filter (most versions).
- Single gas calibration capability.
- Methane elimination switch for environmental applications.
- Security "Adjustment Lockout Switch".
- Confirmation beep (silenceable).
- Meets EPA Method 21 protocol for fugitive emissions testing (most applications).





Authorized Distributor:

33248 Central Ave. Union City, CA 94587
Toll Free: (800) 754-5165 • Phone: (510) 441-5656 • Fax: (510) 441-5650 mail4rki@rkiinstruments.com • www.rkiinstruments.com



Welders Supply Co. P. O. Box 21007 Louisville, Ky. 40221 (502) 635-7531

Certification of Cylinder Content

Thursday, January 27, 2011

Mix Type: Primary Standard

Analytic Accuracy:

± 2%

Serial Number: LL-24367

Cylinder CGA: 350

Test Date:

Approx. PSI: 105

Expiration Date: 27-Jan-14

110127

Analytic Method(s): Gas Chromatography

Cylinder Contents:

Requested Gas

Actual

bal Nitrogen

Balance

85 % Propane

84.44 %

Frank Fogarty Specialty Gas Lab Manger



Welders Supply Co. P. O. Box 21007 Louisville, Ky. 40221 (502) 635-7531

Certification of Cylinder Content

Thursday, January 27, 2011

Mix Type: Primary Standard

Analytic Accuracy:

± 2%

Serial Number: LL-36180

Cylinder CGA: 350

160

Approx. PSI:

Test Date:

110127

Expiration Date: 27-Jan-14

Analytic Method(s): Gas Chromatography

Cylinder Contents:

Requested Gas

Actual

bal Nitrogen

Balance

50 % Propane

50.36 %

Frank Fogarty

Specialty Gas Lab Manger



Welders Supply Co. P. O. Box 21007 Louisville, Ky. 40221-0007 (502) 635-7631

Certification of Cylinder Content

Wonday, October 20, 2008

Primary Standard Mix Type:

±2% Analytic Accuracy:

> LL-36171 Serial Number:

350 Cylinder CGA:

2000 Approx. PSI:

> 081020 Test Date:

20-Oct-11 Expiration Date:

Gravimetric Analytic Method(s):

Gas Chromatography

Cylinder Contents:

Requested Gas

25 % Propane

bal Nitrogen

Actual

24.946 %

Balance

Frank Fogarty

Specialty Gas Lab Manger



LIQUID TECHNOLOGY CORPORATION



"INDUSTRY LEADER IN SPECIALTY GASES"

Certificate of Analysis

- EPA PROTOCOL GAS -

Customer:

Welders Supply Company (Louisville, KY)

Date:

May 24, 2010

Delivery Receipt:

DR-29400

Product:

4.50% Propane/Nitrogen - EPA PROTOCOL

Final Analysis Date:

May 17, 2010

Expiration Date:

May 17, 2013

DO NOT USE BELOW 150 PSIG

Cylinder Data

Cylinder Serial Number:

FF-29457

Cylinder Outlet:

CGA 350

Cylinder Volume: Expiration Date: 30 Cubic Feet May 17, 2013 Cylinder Pressure:

2000 psig, 70°F

Analytical Data

EPA PROTOCOL, Section No. 2.2, Procedure G-1

Replicate Concentrations

Propane: 4.63% +/- 0.046%

Nitrogen: Balance

Reference Standard(s):

SRM/GMIS:

GMIS

Cylinder Number:

CC-70316

Concentration:

3.52% Propane/Nitrogen

Expiration Date:

June 24, 2011

Certification Instrumentation

Component:

Propane

Make/Model:

Agilent 7890A

Serial Number:

CN10736166

Principal of Measurement:

GC-FID

Last Calibration:

April 26, 2010

Analytical uncertainty and NIST Traceability are in compliance with EPA-600/R-97/121.

Certified by:

Mike Duncan

LIQUID TECHNOLOGY CORPORATION

"INDUSTRY LEADER IN SPECIALTY GASES"

Certificate of Analysis

- EPA PROTOCOL GAS -

Customer:

Welders Supply Company (Louisville, KY)

Date:

April 25, 2011

Delivery Receipt:

DR-37004

Product:

2.50% Propane/Nitrogen - EPA PROTOCOL

Final Analysis Date:

April 12, 2011

Expiration Date:

April 12, 2014

DO NOT USE BELOW 150 PSIG

Cylinder Data

Cylinder Serial Number:

FF-24216

Cylinder Outlet: Cylinder Pressure: CGA 350

Cylinder Volume: **Expiration Date:**

30 Cubic Feet April 12, 2014

2000 psig, 70°F

Analytical Data

EPA PROTOCOL, Section No. 2.2, Procedure G-1

Replicate Concentrations Propane: 2.54% +/- 0.025%

Nitrogen: Balance

Reference Standard(s):

SRM/GMIS:

GMIS

Cylinder Number:

CC-116013

Concentration:

2.465% Propane/Nitrogen

Expiration Date:

05/17/12

Certification Instrumentation

Component:

Propane

Make/Model:

Agilent 7890A

Serial Number:

CN10736166

Principal of Measurement:

GC-FID April 01, 2011

Last Calibration:

Analytical uncertainty and NIST Traceability are in compliance with EPA-600/R-97/121.

Certified by:

Adam Strickland



LIQUID TECHNOLOGY CORPORATION

"INDUSTRY LEADER IN SPECIALTY GASES"

Certificate of Analysis

- EPA PROTOCOL GAS -

Customer

Welders Supply Co (Louisville, KY)

Date

March 14, 2011

Delivery Receipt

DR-36591

Gas Standard

1.50% Propane/Nitrogen - EPA PROTOCOL

Final Analysis Date

March 04, 2011

Expiration Date

March 04, 2014

Component

<u>Propane</u>

Balance Gas

Nitrogen

Analytical Data:

DO NOT USE BELOW 150 psig

EPA Protocol, Section No. 2.2, Procedure G-1

Reported Concentrations

Propane: 1.59% +/- 0.01%

Nitrogen: Balance

Reference Standards:

SRM/GMIS:

GMIS

GMIS

Cylinder Number:

CC-166394

CC-116013

Concentration:

9709.18 ppm Propane/Nitrogen

2.465% Propane/Nitrogen

Expiration Date:

07/21/11

05/07/12

Certification Instrumentation

Component:

Propane

Make/Model:

Agilent 7890A

Serial Number:

CN10736166

Principal of Measurement:

GC-FID

Last Calibration:

March 03, 2011

Cylinder Data

Cylinder Serial Number:

FF-34583

Cylinder Outlet:

CGA 350

Cylinder Volume:

29 Cubic Feet

Cylinder Pressure:

1950 psig, 70°F

Expiration Date:

March 14, 2014

Analytical Uncertainty and NIST Traceability are in compliance with EPA-600/R-97/121.

Certified by:

Adam Strickland

Ada Siji

"UNMATCHED EXCELLENCE"

APPENDIX C

COMPUTER PRINTOUT, VOC ANALYZER STRIP CHARTS



Jordan Technologies, Inc. 2820 South English Station Road Louisville, Kentucky 40299 Phone: (502) 267-8344

Fax: (502) 267-8379

Vapor Recovery Unit Performance Test

Test For:

Chevron Products

Port Everglades, Florida

Unit Tested:

Zink VRU

Test Date:

July 14, 2011

Test Personnel:

Tony Fenton

Strip Cart Speed:

150

All data fields are rounded two (2) places following the decimal for display purposes. Internal to the program all data fields are eight (8) digits following the decimal. This will create differences when recalculating the data by hand.

OUTLET CALIBRATION GAS INFORMATION

Outlet analyzer range is 0 - 5%

Allowable range is + or - 5% of actual span gas concentration

Low range span gas concentration Mid range span gas concentration Hi range span gas concentration	1.59 % 2.54 % 4.63 %	Cylinder Cylinder Cylinder	FF34583 FF24216 FF29547
Zero span analyzer reading	0.00 %		
Low range analyzer reading Low range analyzer error	1.62 % 1.89 %		
Mid range analyzer reading Mid range analyzer error	2.52 % -0.79 %		
Hi range analyzer reading Hi range analyzer error	4.61 % -0.43 %		

INLET CALIBRATION GAS INFORMATION

Intlet analyzer range is 0 - 100%

Allowable range is + or - 5% of actual span gas concentration

Low range span gas concentration Mid range span gas concentration	24.946 50.36	%	Cylinder Cylinder	LL36171 LL36180 LL24367
Hi range span gas concentration	84.44	%	Cylinder	LL2430 <i>1</i>
Zero span analyzer reading	0.00	%		
Low range analyzer reading	25.3	%		
Low range analyzer error	1.42	%		
Mid range analyzer reading	50			
Mid range analyzer error	-0.71	%		
Hi range analyzer reading	84.6	%		
Hi range analyzer error	0.19	%		


	FES.	pute p	A 4 ***	gare a de a men	Linia	HCour.	VE	VES	ME	
Time	Baro	Flow P	Atm T	Exh. T	HCin	HCout (Vol %)	(m3)	(m3)	(mg)	
		(mm Hg)	(Deg C)	(Deg C)	(Vol %) 45.70	0.00	23.67	23.04	1687	
7:35:00	759.70	0.14	27.97 28.05	27.93 28.18	43.70	0.02	16.23	15.78	5255	
7:40:00	759.70	0.00 0.00	28.20	27.97	42.60	0.03	6.63	6.45	3563	
7:45:00	759.60	0.00	28.27	27.95	44.10	0.03	17.13	16.68	8636	
7:50:00	759.70	0.12	28.36	28.47	46.20	0.00	19.79	19.23	950	
7:55:00	759.60	0.00	28.49	28.48	44.50	0.02	15.91	15.46	4753	
8:00:00	759.60	0.00	28.61	28.53	41.40	0.02	11.81	11.47	4198	
8:05:00	759.70	0.00	28.79	28.97	35.20	0.00	21.66	21.01	1269	
8:10:00	759.70	0.07		28.98	42.30	0.04	20.98	20.35	14601	
8:15:00	759.80	0.00		29.00	40.90	0.04	10.28	9.97	7298	
8:20:00	759.80	0.00					17.98	17.43	4401	
8:25:00	759.90	0.09				0.00	0.99	0.96	0	
8:30:00	759.90									
8:30:00	Outlet Zero	Check: Read	mg was un	0, WillCiriSi Caubiobioi	n the accep	otable range (otable range ((-3% - 3%)	<i></i>		
8:32:00	Inlet Zero C	neck: Keadir	ig was u.4%	o, WAICH IS I	ti ille accep Aich is in th	ne acceptable	-070 - 070) s renne /2 37	7% - 2 67%)		
8:32:00	Outlet Mid S	ipan Uneck:	Reading wa	ED 40/ wh	MICH IS III U iob io io 650	acceptable	range (2.01 ranga (47% .	. 53%)		
					29.80		0.62	0.61	7271	
8:35:00		0.00					14.64	14.22	0	
8:40:00							12.94	12.51	2175	
8:45:00							8.58	8.30	4328	
8:50:00							11.75	11.36	6008	
8:55:00							18.38	17.69	0	
9:00:00							13.96	13.43	1573	
9:05:00							34.32	33.00	10327	
9:10:00							24.24	23.28	2386	
9:15:00							1.05	1.01	184	
9:20:00							1.36	1.31	240	
9:25:00						otable range			2,0	
9:28:00	Outlet Zero	Check: Real	Dooding w	o, WINCH IS	which is in fl	he acceptable	e ranne (2.3)	7% - 2.67%)		
							19.79	19.04	4042	
9:30:00						ptable range				
9:30:00	Inlet Zero C	neck: Resul	ng was 0.47	70, WIIIUII IS . EO 294 wh	in the accep	acceptable	ranne (47%	- 53%)		
							13.45	12.89	731	
9:35:00							3.00	2.88	543	
9:40:00							15.72	15.06	4245	
9:45:00							16.62	15.86	755	
9:50:00 9:55:00							9.63	9.21	2546	
10:00:00							25.34	24.18	12082	
10:05:00							17.87	17.02	872	
10:00:00							1.44	1.38	0	
10:15:00							24.10	22.98	1808	
10:10:00				_	_		31.18	29.62	2277	
10:25:00							15.35	14.55	2050	
10.20.00						ptable range				
10:29:00						0.02	51.08	48.57	14220	
						ptable range				
10.30.00 10.30.00	Outlet Blid	Snan Cherb	· Reading w	as 2.52% ·	which is in t	he acceptable	le range (2.3	7% - 2.67%)	
10:30:00 Outlet Mid Span Check: Reading was 2.52%, which is in the acceptable range (2.37% - 2.67%) 10:31:00 Inlet Mid Span Check: Reading was 50.1%, which is in the acceptable range (47% - 53%)										
10:31:00								13.21	6238	
10:30:00							6.34	6.02	0	
10:40:00								14.88	327	
10:50:00							11.02	10.46	1915	
10:55:00							49.16	46.47	7995	
, 0.00.00	, 100.10	. 0.0	. • • • • • • • • • • • • • • • • • • •							

11:00:00	760.10	0.00	34.19	36.97	35.20	0.03	7.99	7.55	3758
11:05:00	760.10	0.00	33.32	35.86	37.60	0.03	7.14	6.77	3878
11:10:00	760.10	0.09	35.63	37.17	44.10	0.01	27.18	25.69	6769
11:15:00	760.10	0.00	35.56	38.07	44.70	0.01	16.54	15.58	250 9
11:20:00	760.00	0.00	36.28	36.61	44.20	0.01	1.05	0.99	181
11:25:00	760.00	0.06	37.48	36.47	30.60	0.01	13.88	13.14	1611
11.20.00	Outlet Zero Ch				he acceptal	ble range (-(),15% - 0.15	5%)	
11.21.00	Inlet Zero Che	ck: Reading v	was 0.5%.	which is in I	the accepta	ble range (-	3% - 3%)		
11.20.00	Outlet Mid Spa	on Check: Re	ading was	2.52%. whi	ch is in the	acceptable	range (2.37 ⁹	% - 2.67%)	
11.20.00	Inlet Mid Span	Check: Read	ding was 5	0.3%, which	is in the ac	ceptable ra	nge (47% -	53%)	
11:30:00	760.00	0.00	37.50	38.20	35.90	0.01	8.75	8.24	1116
11:35:00	760.00	0.00	36.46	37.68	34.00	0.01	11.24	10.60	2600
11:40:00	760.00	0.07	35.50	37.82	38.00	0.03	17.27	16.28	7987
11:45:00	760.00	0.02	36.79	38.38	45.50	0.00	13.22	12.44	0
11:50:00	760.00	0.00	38.34	36.66	41.30	0.00	0.20	0.19	0
11:55:00	760.00	0.00	36.60	35.97	36.00	0.00	1.81	1.72	0
12:00:00	760.00	0.06	38.08	38.10	17.50	0.00	20.10	18.94	0
12:05:00	760.00	0.00	38.25	38.22	31.60	0.00	4.93	4.64	68
12:10:00	760.00	0.02	35.03	37.95	38.40	0.02	28.32	26.68	9766
12:15:00	760.00	0.02	34.28	37.51	39.70	0.01	18.60	17.56	4723
12:10:00	760.00	0.00	38.22	39.46	45.30	0.00	12.74	11.95	0
	760.00	0.00	34.49	38.61	38.10	0.00	14.92	14.03	360
12:25:00	Outlet Zero Cl							5%)	
	759.90	0.15	33.09	37.76	5.20	0.06	24.13	22.75	24854
12:30:00	Inlet Zero Che								
42.20.00	Outlet Mid Sp	es Chook: D	ved 0.2 /0,	2 51% wh	ich is in the	acceptable	range (2.37	% - 2.67%)	
42.34.00	Inlet Mid Spar	an Chook: Bos	dina wae i	:0 1% which	h is in the a	cceptable ra	ange (47% -	53%)	
	759.90	0.01	32.76	37.76	28.10	0.00	12.63	11.91	436
12:35:00	760.00	0.00	34.31	37.78	7.90	0.00	18.97	17.89	491
12:40:00	760.00	0.00	34.15	37.82	21.30	0.01	13.17	12.41	2272
12:45:00 12:50:00	760.00	0.12	32.63	37.30	45.30	0.00	23.93	22.60	538
12:55:00	760.00	0.00	32.89	36.55	48.00	0.00	3.74	3.54	0
13:00:00	760.00	0.00	36.42	35.77	42.70	0.00	1.10	1.05	33
13:05:00	760.00	0.05	37.02	37.09	20.80	0.00	17.78	16.80	1353
13:00:00	760.00	0.00	37.05	37.98	16.40	0.00	3.45	3.26	0
13:15:00	760.00	0.00	36.93	36.08	17.60	0.00	0.79	0.75	0
13:10:00		0.10	37.03	36.64	34.10	0.00	14.16	13.40	834
13.20.00	Outlet Mid Sp							7% - 2.67%)	
13:25:00		0.00	37.14	39.77	45.10	0.00	17.67	16.55	0
13:30:00		0.00	36.63	37.88	44.30	0.00	0.57	0.53	9
13.30.00								15%)	
13:30:00 Outlet Zero Check: Reading was 0%, which is in the acceptable range (-0.15% - 0.15%) 13:31:00 Inlet Zero Check: Reading was 0.4%, which is in the acceptable range (-3% - 3%)									
42-22-00	Inlet Mid Spa	n Check: Re:	adina was	50%. which	is in the ac	ceptable rar	ige (47% - 5	53%)	
13:32:00		,, onest, ite				•	- '		
10.00.00									
SUMS							1043.50	995.26	229892.85
AVRGS	760.02	0.06	32.84	34.20	33.69	0.02			
AVINOU	, 00.02	U, WW	·						


PRELIMINARY TEST RESULTS

There were 72 test intervals, 72 of which had flow


Average Barometric Pressure was Average Flow Pressure was Average Ambient Temerature was Average Inlet Concentration was Average Outlet Concentration was	0.0 6 32.84 33.69	mm Hg mm Hg Deg C Vol. % Vol. %
Total volume emitted was	1044	cubic meters
Total Volume Emitted standardized was	995	cubic meters
Total milligrams emitted was	229893	mg
Accountable gallons loaded was	216,100	gallons
Total gallons loaded was	267,308	gallons
Accountable liters loaded was	817,939	liters
Total Liters loaded was	1,011,761	liters
Accountable milligrams emitted per liter loaded was	0.28	mg/L
Total milligrams emitted per liter loaded was	0.23	mg/L
Unit Efficiency accountable was	99.95	%
Unit Efficiency total was	99.96	%



CHEVRON PRODUCTS - VRU TEST - PORT EVERGLADES, FL - 7/14/11 - PAGE 2



CHEVRON PRODUCTS - VRU TEST - PORT EVERGLADES, FL - 7/14/11 - PAGE 3



CHEVRON PRODUCTS - VRU TEST - PORT EVERGLADES, FL - 7/14/11 - PAGE 4

