

Aux Boilers CO/L CGA

Indiantown CoGen

Test Information

Test Date: 3/2/2012
Facility: Indiantown CoGen
Unit: Aux Boilers

Analyzer Information

Range: Low
Instrument Span: 200 ppm
Manufacturer: Siemens
Model: UltramatOxymat6
Serial Number: 7MB20231EA20-0BD1

Run Number	Time	Reference Gas	CEMS Response	
Low Gas				
1	8:58 AM	50.700	51.030	Allowable Reference Values: 40-60 ppm (20-30% of span)
2	9:08 AM	50.700	50.630	
3	9:18 AM	50.700	50.670	
Mean (ppm)		50.700	50.777	
Mean Difference (ppm)		0.077	Limit: 5	Passed
Linearity Error		0%	Limit: 15%	Passed
Mid Gas				
1	9:03 AM	108.000	107.910	Allowable Reference Values: 100-120 ppm (50-60% of span)
2	9:13 AM	108.000	108.070	
3	9:23 AM	108.000	107.940	
Mean (ppm)		108.000	107.973	
Mean Difference (ppm)		0.027	Limit: 5	Passed
Linearity Error		0%	Limit: 15%	Passed

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses

Aux Boilers NOx CGA

Indiantown CoGen

Test Information

Test Date: 3/2/2012
Facility: Indiantown CoGen
Unit: Aux Boilers

Analyzer Information

Range: Single
Instrument Span: 300 ppm
Manufacturer: Thermo Environmental
Model: 42i
Serial Number: 0617417188

Run Number	Time	Reference Gas	CEMS Response	
Low Gas				
1	8:58 AM	72.900	74.390	Allowable Reference Values: 60-90 ppm (20-30% of span)
2	9:08 AM	72.900	74.540	
3	9:18 AM	72.900	74.610	
Mean (ppm)		72.900	74.513	
Mean Difference (ppm)		1.613	Limit: 5	Passed
Linearity Error		2%	Limit: 15%	Passed
Mid Gas				
1	9:03 AM	165.000	168.790	Allowable Reference Values: 150-180 ppm (50-60% of span)
2	9:13 AM	165.000	168.800	
3	9:23 AM	165.000	168.460	
Mean (ppm)		165.000	168.683	
Mean Difference (ppm)		3.683	Limit: 5	Passed
Linearity Error		2%	Limit: 15%	Passed

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses

Aux Boilers CO/H CGA

Indiantown CoGen

Test Information

Test Date: 3/2/2012
Facility: Indiantown CoGen
Unit: Aux Boilers

Analyzer Information

Range: High
Instrument Span: 1000 ppm
Manufacturer: Siemens
Model: UltramatOxymat6
Serial Number: 7MB20231EA20-0BD1

Run Number	Time	Reference Gas	CEMS Response	
Low Gas				
1	10:23 AM	251.000	251.200	Allowable Reference Values: 200-300 ppm (20-30% of span)
2	10:33 AM	251.000	251.900	
3	10:43 AM	251.000	251.900	
	Mean (ppm)	251.000	251.667	
	Mean Difference (ppm)	0.667	Limit: 5	Passed
	Linearity Error	0%	Limit: 15%	Passed
Mid Gas				
1	10:28 AM	549.000	551.300	Allowable Reference Values: 500-600 ppm (50-60% of span)
2	10:38 AM	549.000	551.300	
3	10:48 AM	549.000	551.900	
	Mean (ppm)	549.000	551.500	
	Mean Difference (ppm)	2.500	Limit: 5	Passed
	Linearity Error	0%	Limit: 15%	Passed

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses

Aux Boilers O2 CGA

Indiantown CoGen

Test Information

Test Date: 3/2/2012
Facility: Indiantown CoGen
Unit: Aux Boilers

Analyzer Information

Range: Single
Instrument Span: 20 %O2
Manufacturer: California Analytical
Model: UltramatOxymat6
Serial Number: 7MB20231EA20-0BD1

Run Number	Time	Reference Gas	CEMS Response
------------	------	---------------	---------------

Low Gas

1	10:23 AM	5.040	5.000	Allowable Reference Values: 4-6 %O2
2	10:33 AM	5.040	5.000	
3	10:43 AM	5.040	5.000	

Mean (%O2)	5.040	5.000	
Mean Difference (%O2)	0.040	Limit: N/A	
Linearity Error	1%	Limit: 15%	Passed

Mid Gas

1	10:28 AM	10.020	10.050	Allowable Reference Values: 8-12 %O2
2	10:38 AM	10.020	10.040	
3	10:48 AM	10.020	10.040	

Mean (%O2)	10.020	10.043	
Mean Difference (%O2)	0.023	Limit: N/A	
Linearity Error	0%	Limit: 15%	Passed

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses