

Indiantown Cogeneration, L.P.

PO Box 1799
13303 SW Silver Fox Lane
Indiantown, FL 34956
P: (772) 597-6500
F: (772) 597-6210

April 1st, 2013

RECEIVED

APR 02 2013

DIVISION OF AIR
RESOURCE MANAGEMENT

Mr. Errin Pichard
Florida Department of Environmental Protection
Division of Air Resource Management
2600 Blair Stone Road MS 5500.
Tallahassee, FL 33401

Re: Indiantown Cogeneration L.P. First Quarter of 2013 Linearity, Opacity Audit & Cylinder Gas Audit Report {Permit Nos: 0850102 (PSD-FL-168)}

Dear Mr. Pichard

Attached please find the First Quarter of 2013 Linearity/Calibration Gas Audits (CGA), Opacity Audits and summary of the audit results for Indiantown Cogeneration, L.P. (ICLP) Main Boiler CEMDAS and Inlet reactors A & B. These audits were performed following the guidelines established in 40CFR60/75, Appendix B.

Summary of Cylinder Gas Audit Results

Parameter	Low Gas Error	Mid Gas Error	High Gas Error	Limit
SO2 Inlet A	4.00	3.00		15
CO2 Inlet A	4.00	3.00		15
SO2 Inlet B	2.00	1.00		15
CO2 Inlet B	2.00	2.00		15
NOX Aux	N/A	N/A		15
O2 Aux	N/A	N/A		15
CO Lo Aux	N/A	N/A		15
CO Hi Aux	N/A	N/A		15
NOX Main	2.80	0.30	0.30	5.0
SO2 Main	4.30	1.30	2.70	5.0
CO2 Main	0.45*	2.50	3.00	5.0
Opacity Cal Error	0.30	0.30	0.60	3.0

- * Alternate method passed at 0.45 ppm

Linearity Audit (Main Boiler)

- The quarterly Linearity Audit for the First Quarter 2013 was completed on March 7th, 2013 for the Main Boiler's sulfur dioxide, nitrogen oxides, and carbon dioxide analyzers. All were within tolerances.

Cylinder Gas Audits Reactor A & B

- The quarterly CGA's for the First Quarter 2013 was completed on March 7th, 2013 for Reactors A & B Sulfur dioxide and carbon dioxide. All were found within tolerances.

Cylinder Gas Audits Auxiliary Boilers


This is a RATA Quarter.

Opacity Audit (Main Boiler)

- The opacity meter was tested on March 8th, 2013 and found to be within the 3% tolerance in all ranges.

Attached are the Audit reports for your review. In accordance to Chapter 62-213-440(1)(b)3-c, F.A.C., I certify that this document and all attachments were prepared under my direction or supervision in accordance with a system designed to assure that qualified personnel properly gather and evaluate the information submitted. Should you have any questions please contact Nicholas Laryea at (772) 597-6500 x 19.

Sincerely,


Todd Shirley
Projects General Manager

cc: Lee Hoefert (DEP South East District)
Kellee Fletcher
Nicholas Laryea
File # 2.4.1

Main Boiler Linearity

LINEARITY PERFORMANCE AUDIT

FOR

Indiantown Cogeneration

Indiantown, FL

Units: Main Emission unit E001

DILUTION EXTRACTIVE

First Quarter 2013

PREPARED BY:

Indiantown Cogeneration

Section 1 Introduction

Facility Location: Indiantown Cogeneration, L.P. operates the Indiantown Cogeneration Plant, which is located in Martin County at 3303 SW Sliver Fox Lane, Indiantown, Florida.

The Indiantown Cogeneration Plant is a cogeneration facility which generates electricity for sale and exports steam to the Louis Dreyfus Citrus Processing Plant. The Facility includes one high-pressure pulverized coal main boiler (PC boiler) rated at 3,422 million British thermal units (MMBtu)/hour heat input, and has a nominal net electrical power output of approximately 330 megawatts (MW). It is permitted to fire natural gas, propane, or No. 2 fuel oil for startup, shutdown, or load changes.

Also included are two natural gas (or propane) fired identical auxiliary boilers used for supplying steam to the steam host during times when the PC boiler is offline, as well as during PC boiler startup and shutdown periods. The two identically sized packaged water-tube steam boilers have a combined rated maximum capacity of 350 MMBtu/hr.

Steam produced by the auxiliary boilers is not used to generate electricity. In addition, the facility has a variety of ancillary equipment needed to support operations as a coal-fired cogeneration plant.

Indiantown Cogeneration of Indiantown, Florida conducted this quarter's linearity test audit at Indiantown Cogeneration on March 07 2013 by Eric Rammel. The audit of the Continuous Emission Monitoring System was conducted for NOx, SO2 and CO2.

Our assessment of these results indicates that all of the analyzers evaluated during this test program meet the criteria for the linearity test requirements for 40 CFR 75 Section 3. Support Data summarizes the results for the Linearity Test Audit

Reviewed by:

Nicholas Rangel

Date:

3/20/2013

TABLE OF CONTENTS

<i>Section 1 Introduction.....</i>	<i>3</i>
<i>Section 2 Linearity Test Audit Procedures</i>	<i>4</i>
<i>Section 3 Support Data Section 4 Gas Cylinder Certificates</i>	

Summary of Cylinder Gas Audit Results

Parameter	Low Gas	Mid Gas	High Gas
CO2	.45	2.5	3.0
SO2	4.3	1.3	2.7
NOx	2.8	.3	.3

Section 2 Linearity Test Audit Procedures

Each Continuous Emission Monitor (CEM) must be audited four out of four calendar quarters of the year, and anytime there is a major repair. As a part of the Quality Control (QC) and Quality Assurance (QA) procedures, the quality of the data produced is evaluated by response accuracy compared to known standards.

This quarter's linearity was conducted in accordance with the QA/QC procedure outlined in 40 CFR 75, Appendix A.

The audit consisted of challenging the CEM with an audit gas of known concentration with 3 upscale levels of gas; Low level concentration: 20-30% of span, Mid level concentration: 50-60% of span, High level concentration: 80-100% of span. The audit cylinders contain pollutant or diluent gas certified in accordance with U.S. EPA protocol 1.

The audit gases were introduced into the entire sampling and analysis system through the normal part of the daily QC gases.

The procedure was conducted as follows:

- 1 Audit cylinder 1 was connected to the system.
- 2 Manual span was initiated for approximately 15 minutes until a stable response was achieved.
- 3 Values were recorded as the system was allowed to operate in a normal sampling and analysis manner without adjustment
- 4 The first audit cylinder was removed and replaced by audit cylinder 2.
- 5 Manual span was initiated for approximately 15 minutes until a stable response was achieved.
- 6 The second audit cylinder was removed and replaced by audit cylinder 3.
- 7 Manual span was initiated for approximately 15 minutes until a stable response was achieved.
- 8 This series of steps was repeated through three audit runs.

For each audit cylinder (or audit point), the error in linearity for each calibration (low, mid, and high levels) was determined. The average of the linearity at each level was determined for SO₂ and NO_x pollutant concentration monitors by the following equations:

$$LE = \frac{|R-A|}{R} \times 100 \text{ or } |R - A| \leq 5 \text{ ppm}$$

For CO₂ or O₂ monitors:

$$LE = |R-A| \leq 5\% \text{ or } |R - A| \leq 0.5\% \text{ CO}_2 \text{ or O}_2$$

which ever is less restrictive per audit level

LE = Linearity Error $LE \leq 5\%$

R = Reference Value for High, Mid, or Low Range Gas A = Average Response of the Monitors Value

Section 3 Support Data

Main Boiler NOx Linearity

Indiantown CoGen

Test Information

Test Date: 3/7/2013
 Facility: Indiantown CoGen
 Unit: Main Boiler
 ORIS: 050976
 Test Reason(s):
 Aborted: No

Analyzer Information

Range: Single
 Instrument Span: 250 ppm
 Component ID: 110
 Monitoring System ID: N10
 Manufacturer: Thermo Environmental Inst
 Model: 42I
 Serial Number: CM072660054

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information	
Low Gas					
1	12:48 PM	64.100	65.880	Allowable Reference Values: 50-75 ppm (20-30% of span) Cylinder ID: CC185304 Expiration date: 8/30/2013 EPA vendor ID: E12011 Cylinder contains: CO2,NO,SO2,BALN	
2	1:18 PM	64.100	65.890		
3	1:48 PM	64.100	65.870		
Mean (ppm)		64.100	65.880		
Mean Difference (ppm)		1.780	Limit: 5 Passed		
Linearity Error		2.8%	Limit: 5.0% Passed		
Mid Gas					
1	12:58 PM	131.000	131.330		Allowable Reference Values: 125-150 ppm (50-60% of span) Cylinder ID: CC-166604 Expiration date: 4/23/2014 EPA vendor ID: E12012 Cylinder contains: CO2,NO,SO2,BALN
2	1:28 PM	131.000	131.410		
3	1:58 PM	131.000	131.280		
Mean (ppm)		131.000	131.340		
Mean Difference (ppm)		0.340	Limit: 5 Passed		
Linearity Error		0.3%	Limit: 5.0% Passed		
High Gas					
1	1:08 PM	214.000	213.430	Allowable Reference Values: 200-250 ppm (80-100% of span) Cylinder ID: CC56344 Expiration date: 2/18/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN	
2	1:38 PM	214.000	213.990		
3	2:08 PM	214.000	212.940		
Mean (ppm)		214.000	213.453		
Mean Difference (ppm)		0.547	Limit: 5 Passed		
Linearity Error		0.3%	Limit: 5.0% Passed		

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses

Main Boiler SO2 Linearity

Indiantown CoGen

Test Information

Test Date: 3/7/2013
Facility: Indiantown CoGen
Unit: Main Boiler
ORIS: 050976
Test Reason(s):
Aborted: No

Analyzer Information

Range: Single
Instrument Span: 140 ppm
Component ID: 160
Monitoring System ID: S10
Manufacturer: Thermo Environmental Inst
Model: 43I
Serial Number: CM07260055

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information	
Low Gas					
1	12:48 PM	33.800	35.070	Allowable Reference Values: 28-42 ppm (20-30% of span) Cylinder ID: CC185304 Expiration date: 8/30/2013 EPA vendor ID: E12011 Cylinder contains: CO2,NO,SO2,BALN	
2	1:18 PM	33.800	35.350		
3	1:48 PM	33.800	35.340		
Mean (ppm)		33.800	35.253		
Mean Difference (ppm)		1.453	Limit: 5 Passed		
Linearity Error		4.3%	Limit: 5.0% Passed		
Mid Gas					
1	12:58 PM	75.200	75.940		
2	1:28 PM	75.200	76.290		
3	1:58 PM	75.200	76.270		
Mean (ppm)		75.200	76.167		
Mean Difference (ppm)		0.967	Limit: 5 Passed		
Linearity Error		1.3%	Limit: 5.0% Passed		
High Gas					
1	1:08 PM	129.000	124.510	Allowable Reference Values: 112-140 ppm (80-100% of span) Cylinder ID: CC56344 Expiration date: 2/18/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN	
2	1:38 PM	129.000	126.150		
3	2:08 PM	129.000	125.830		
Mean (ppm)		129.000	125.497		
Mean Difference (ppm)		3.503	Limit: 5 Passed		
Linearity Error		2.7%	Limit: 5.0% Passed		

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses

Main Boiler CO2 Linearity

Indiantown CoGen

Test Information

Test Date: 3/7/2013
 Facility: Indiantown CoGen
 Unit: Main Boiler
 ORIS: 050976
 Test Reason(s):
 Aborted: No

Analyzer Information


Range: Single
 Instrument Span: 20 %CO2
 Component ID: 140
 Monitoring System ID: C10
 Manufacturer: California Analytical
 Model: ZRH-1
 Serial Number: A7B3749T

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information
Low Gas				
1	12:48 PM	4.940	5.370	Allowable Reference Values: 4-6 %CO2 (20-30% of span) Cylinder ID: CC185304 Expiration date: 8/30/2013 EPA vendor ID: E12011 Cylinder contains: CO2,NO,SO2,BALN
2	1:18 PM	4.940	5.400	
3	1:48 PM	4.940	5.410	
Mean (%CO2)		4.940	5.393	
Mean Difference (%CO2)		0.453	Limit: 0.5 Passed	
Linearity Error		9.2%	Limit: 5.0%	
Mid Gas				
1	12:58 PM	10.800	11.050	Allowable Reference Values: 10-12 %CO2 (50-60% of span) Cylinder ID: CC-166604 Expiration date: 4/23/2014 EPA vendor ID: E12012 Cylinder contains: CO2,NO,SO2,BALN
2	1:28 PM	10.800	11.070	
3	1:58 PM	10.800	11.100	
Mean (%CO2)		10.800	11.073	
Mean Difference (%CO2)		0.273	Limit: 0.5 Passed	
Linearity Error		2.5%	Limit: 5.0% Passed	
High Gas				
1	1:08 PM	18.100	17.530	Allowable Reference Values: 16-20 %CO2 (80-100% of span) Cylinder ID: CC56344 Expiration date: 2/18/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN
2	1:38 PM	18.100	17.500	
3	2:08 PM	18.100	17.640	
Mean (%CO2)		18.100	17.557	
Mean Difference (%CO2)		0.543	Limit: 0.5	
Linearity Error		3.0%	Limit: 5.0% Passed	

Linearity Error (LE) Determination: $LE = (|R-A| / R) * 100$

R = Reference gas value

A = Mean of actual CEMS responses



LIQUID TECHNOLOGY CORPORATION

"INDUSTRY LEADER IN SPECIALTY GASES"

Main. CEMS

Certificate of Analysis

- EPA PROTOCOL GAS -

IN Service

10-19-11

Customer Indiantown Cogen, LP (Indiantown, FL)
Date August 30, 2011
Delivery Receipt DR-38427
Gas Standard 63.0 ppm NO, 35.0 ppm SO2, 5.00% CO2/Nitrogen - EPA PROTOCOL
Final Analysis Date August 30, 2011
Expiration Date August 30, 2013

DO NOT USE BELOW 150 psig

Analytical Data:

EPA Protocol, Section No. 2.2, Procedure G-1.

Reported Concentrations

Nitric Oxide: 64.1 ppm +/- 0.64 ppm

Sulfur Dioxide: 33.8 ppm +/- 0.33 ppm

Carbon Dioxide: 4.94% +/- 0.04%

Nitrogen: Balance

Total NOx: 64.1 ppm

**** Total NOx for Reference Use Only ****

linearity

Reference Standards

SRM/GMIS:	GMIS/GMIS	GMIS/GMIS	GMIS
Cylinder Number:	CC-233272/EB-0017545	CC-233316/CC-233274	CC-159026
Concentration:	49.23 ppm NO/96.85 ppm NO	10.67 ppm/53.31 ppm SO2	4.974% CO2
Expiration Date:	04/14/13 - 04/25/13	05/03/12 - 09/20/12	10/14/12

Certification Instrumentation

Component:	Nitric Oxide	Sulfur Dioxide	Carbon Dioxide
Make/Model:	Nicolet - NEXUS 470	Nicolet - NEXUS 470	Nicolet - NEXUS 470
Serial Number:	AEP9900154	AEP9900154	AEP9900154
Principal of Measurement:	FTIR	FTIR	FTIR
Last Calibration:	August 10, 2011	August 10, 2011	August 10, 2011

Cylinder Data

Cylinder Number:	CC-185304	Cylinder Volume:	133 Cubic Feet
Cylinder Outlet:	CGA 660	Cylinder Pressure:	1900 psig, 70°F


Analytical Uncertainty and NIST Traceability are in compliance with EPA-600/R-97/121.

Certified by:

Adam Strickland

PGVP Vendor ID: E12011

"UNMATCHED EXCELLENCE"



LIQUID TECHNOLOGY CORPORATION

"INDUSTRY LEADER IN SPECIALTY GASES"

Certificate of Analysis - EPA PROTOCOL GAS -

In Service

5-11-12

Main

Customer Indiantown Cogen, LP (Indiantown, FL)
Date April 24, 2012
Delivery Receipt DR-42043
Gas Standard 130 ppm NO, 76.0 ppm SO₂, 11.0% CO₂/Nitrogen - EPA PROTOCOL
Final Analysis Date April 23, 2012
Expiration Date April 23, 2014

DO NOT USE BELOW 150 psig

Analytical Data:
EPA Protocol, Section No. 2.2, Procedure G-1.

Reported Concentrations
Nitric Oxide: 131 ppm +/- 1.3 ppm
Sulfur Dioxide: 75.2 ppm +/- 0.75 ppm
Carbon Dioxide: 10.8% +/- 0.10%
Nitrogen: Balance
Total NOx: 131 ppm

** Total NOx for Reference Use Only **

Reference Standards

SRM/GMIS:	GMIS/GMIS	GMIS/GMIS	GMIS
Cylinder Number:	EB-0023236/CC-88793	EB-0014653	CC-165377
Concentration:	96.27 ppm NO/262.42 ppm NO	103.89 ppm SO ₂	9.924% CO ₂
Expiration Date:	01/12/14 - 01/13/14	12/01/12	10/14/12

Certification Instrumentation

Component:	Nitric Oxide	Sulfur Dioxide	Carbon Dioxide
Make/Model:	Nicolet 6700	Nicolet 6700	Nicolet 6700
Serial Number:	APW1100563	APW1100563	APW1100563
Principal of Measurement:	FTIR	FTIR	FTIR
Last Calibration:	April 06, 2012	April 06, 2012	April 06, 2012

Cylinder Data

Cylinder Number:	CC-166604	Cylinder Volume:	133 Cubic Feet
Cylinder Outlet:	CGA 660	Cylinder Pressure:	1900 psig, 70°F


Analytical Uncertainty and NIST Traceability are in compliance with EPA-600/R-97/121.

Certified by:

Adam Strickland

PGVP Vendor ID: E12012

"UNMATCHED EXCELLENCE"



LIQUID TECHNOLOGY CORPORATION
"INDUSTRY LEADER IN SPECIALTY GASES"

3/7/13

Certificate of Analysis
- EPA PROTOCOL GAS -

DAILY SPAN

Customer Indiantown Cogen, LP (Indiantown, FL)
Date February 19, 2013
Delivery Receipt DR-45823
Gas Standard 222 ppm NO, 131 ppm SO2, 18.0% CO2/Nitrogen - EPA PROTOCOL
Final Analysis Date February 18, 2013
Expiration Date February 18, 2021

DO NOT USE BELOW 100 psig

Analytical Data:
EPA Protocol, Section No. 2.2, Procedure G-1.

Reported Concentrations
Nitric Oxide: 214 ppm +/- 0.32 ppm
Sulfur Dioxide: 129 ppm +/- 0.89 ppm
Carbon Dioxide: 18.1% +/- 0.06%
Nitrogen: Balance
Total NOx: 216 ppm

** Total NOx for Reference Use Only **

Reference Standards

SRM/GMIS:	GMIS/GMIS	GMIS/GMIS	GMIS
Cylinder Number:	ND-45697/ND-45515	CC-54548/CC-251490	CC-184404
Concentration:	96.46 ppm NO/246.26 ppm NO	102.43 ppm SO2/507.87 ppm SO2	19.87% CO2
Expiration Date:	01/12/14 - 08/23/20	04/12/13 - 04/12/13	02/04/13

Certification Instrumentation

Component:	Nitric Oxide	Sulfur Dioxide	Carbon Dioxide
Make/Model:	Nicolet 6700	Nicolet 6700	Nicolet 6700
Serial Number:	APW1100563	APW1100563	APW1100563
Principal of Measurement:	FTIR	FTIR	FTIR
Last Calibration:	January 30, 2013	January 30, 2013	January 30, 2013

Cylinder Data

Cylinder Number:	CC-56344	Cylinder Volume:	131 Cubic Feet
Cylinder Outlet:	CGA 660	Cylinder Pressure:	1875 psig, 70°F

Analytical Uncertainty and NIST Traceability are in compliance with EPA-600/R-12/531.

Certified by:

David Scott

PGVP Vendor ID: E12013

"UNMATCHED EXCELLENCE"

Reactors A & B
CGA

CYLINDER GAS AUDIT

FOR

Indiantown Cogeneration

Indiantown, FL

**Unit: Reactor Emission Unit ID Nos. 001
Inlet A and B
DILUTION EXTRACTIVE**

First Quarter 2013

PREPARED BY:

Indiantown Cogeneration

TABLE OF CONTENTS

<i>Section 1</i>	<i>Introduction</i>
<i>Section 2</i>	<i>Cylinder Gas Audit Procedures</i>
<i>Section 3</i>	<i>Cylinder Gas Audit Data Sheets</i>
<i>Section 4</i>	<i>Cylinder Gas Certification Sheets</i>

LIST OF TABLES

Table	Page
Table 1.1: Summary of Cylinder Gas Audit Results	2

Section 1 Introduction

Facility Location: Indiantown Cogeneration, L.P. operates the Indiantown Cogeneration Plant, which is located in Martin County at 3303 SW Sliver Fox Lane, Indiantown, Florida.

The Indiantown Cogeneration Plant is a cogeneration facility which generates electricity for sale and exports steam to the Louis Dreyfus Citrus Processing Plant. The Facility includes one high-pressure pulverized coal main boiler (PC boiler) rated at 3,422 million British thermal units (MMBtu)/hour heat input, and has a nominal net electrical power output of approximately 330 megawatts (MW). It is permitted to fire natural gas, propane, or No. 2 fuel oil for startup, shutdown, or load changes.

Also included are two natural gas (or propane) fired identical auxiliary boilers used for supplying steam to the steam host during times when the PC boiler is offline, as well as during PC boiler startup and shutdown periods. The two identically sized packaged water-tube steam boilers have a combined rated maximum capacity of 350 MMBtu/hr.

Steam produced by the auxiliary boilers is not used to generate electricity. In addition, the facility has a variety of ancillary equipment needed to support operations as a coal-fired cogeneration plant.

Indiantown Cogenerating, of Indiantown, Florida conducted a Cylinder Gas Audit at Indiantown Cogeneration in Indiantown, FL on March 6 & 7, 2012 conducted by Eric Rammel.

The audit of the Continuous Emission Monitoring System was conducted for SO₂ & CO₂. Our assessment of this quarter CGA results indicates that all of the analyzers evaluated during this test program meet the accuracy requirements as outlined in 40 CFR 60, Appendix F. Table 1.1 summarizes the results for the cylinder gas audit.

Reviewed by: Michael Langes

Date: 3/20/2013

Section 2

Cylinder Gas Audit Procedures

Each Continuous Emission Monitor (CEM) must be audited three out of four calendar quarters of each year. As a part of the Quality Control (QC) and Quality Assurance (QA) procedures, the quality of data produced is evaluated by response accuracy compared to known standards.

The CGA for this quarter was conducted in accordance with the QA/QC procedure outlined in 40 CFR 60, Appendix F.

The Audit consisted of challenging the CEM with an audit gas of known concentration with 2 upscale levels of gas, at 20-30% of the system span and at 50-60% of the system span. The audit cylinders contain pollutant or diluent gas certified in accordance with U.S. EPA protocol 1.

The audit gases were introduced into the entire sampling and analysis system through the normal part of the daily QC gases.

The procedure was conducted as follows:

1. Audit cylinder 1 was connected to the system.
2. Manual span was initiated until a stable response was achieved.
3. Values were recorded as the system was allowed to operate in a normal sampling and analysis manner without adjustment.
4. The first audit cylinder was removed and replaced by audit cylinder 2.
5. Manual span was initiated for approximately 15 minutes until a stable response was achieved.
6. This series of steps was repeated through three audit runs.

Cylinder Gas Audit Reactor A and B Inlet E.U. ID Nos. 001

For each audit cylinder (or audit point), the percent accuracy was determined. The average of the accuracy was determined by the following equation:

$$A = \frac{(Cm - Ca) \times 100}{Ca}$$

Where:

- A = Accuracy of CEMS (%)
Cm = Average CEMS response during audit in applicable standard or concentration (ppm or %)
Ca = Average audit (cylinder gas certified value) in units of applicable standard or concentration (ppm or %)

Accuracy (A) value of $\pm 15\%$ or less is considered acceptable for criteria pollutants or diluent gas. 40 CFR 60 appendix F 5.3.2. (2).

Section 3
Cylinder Gas Audit Data Sheets

Reactor A CO2 CGA

Indiantown CoGen

Test Information

Test Date: 3/6/2013
 Facility: Indiantown CoGen
 Unit: Reactor A

Analyzer Information

Range: Single
 Instrument Span: 20 %CO2
 Manufacturer: California Analytical
 Model: ZRH1 DFY2-6BAYY
 Serial Number: A9E3765T

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information
Low Gas				
1	2:10 PM	6.490	6.780	Allowable Reference Values: 5-8 %CO2 Cylinder ID: CC6414 Expiration date: 7/6/2013 EPA vendor ID: F12012 Cylinder contains: CO2,NO,SO2,BALN
2	2:32 PM	6.490	6.780	
3	2:54 PM	6.490	6.750	
Mean (%CO2)		6.490	6.770	
Mean Difference (%CO2)		0.280	Limit: N/A	
CEMS Accuracy		4%	Limit: 15%	Passed
Mid Gas				
1	2:21 PM	11.800	12.220	Allowable Reference Values: 10-14 %CO2 Cylinder ID: EB-0014601 Expiration date: 2/12/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN
2	2:43 PM	11.800	12.200	
3	3:05 PM	11.800	12.200	
Mean (%CO2)		11.800	12.207	
Mean Difference (%CO2)		0.407	Limit: N/A	
CEMS Accuracy		3%	Limit: 15%	Passed

Linearity Error (CEMS Accuracy) Determination (A): $A = (|C_m - C_a| / C_a) * 100$

C_a = Reference gas value

C_m = Mean of actual CEMS responses

Reactor A SO2 CGA

Indiantown CoGen

Test Information

Test Date: 3/6/2013
Facility: Indiantown CoGen
Unit: Reactor A

Analyzer Information

Range: Single
Instrument Span: 2000 ppm
Manufacturer: Thermo Environmental
Model: 43C
Serial Number: 43C-64046-342

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information
Low Gas				Allowable Reference Values: 400-600 ppm (20-30% of span) Cylinder ID: CC6414 Expiration date: 7/6/2013 EPA vendor ID: F12012 Cylinder contains: CO2,NO,SO2,BALN
1	2:10 PM	517.300	539.800	
2	2:32 PM	517.300	539.800	
3	2:54 PM	517.300	537.500	
Mean (ppm)		517.300	539.033	
Mean Difference (ppm)		21.733	Limit: 5	
CEMS Accuracy		4%	Limit: 15% Passed	
Mid Gas				Allowable Reference Values: 1000-1200 ppm (50-60% of span) Cylinder ID: EB-0014601 Expiration date: 2/12/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN
1	2:21 PM	1140.000	1171.000	
2	2:43 PM	1140.000	1171.900	
3	3:05 PM	1140.000	1170.400	
Mean (ppm)		1140.000	1171.100	
Mean Difference (ppm)		31.100	Limit: 5	
CEMS Accuracy		3%	Limit: 15% Passed	

Linearity Error (CEMS Accuracy) Determination (A): $A = (|C_m - C_a| / C_a) * 100$

C_a = Reference gas value

C_m = Mean of actual CEMS responses

Reactor B CO2 CGA

Indiantown CoGen

Test Information

Test Date: 3/7/2013
 Facility: Indiantown CoGen
 Unit: Reactor B

Analyzer Information

Range: Single
 Instrument Span: 20 %CO2
 Manufacturer: California Analytical
 Model: ZRH1 DFY2-6BAYY
 Serial Number: A9E3767T

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information
Low Gas				
1	3:18 PM	6.490	6.630	Allowable Reference Values: 5-8 %CO2 Cylinder ID: CC6414 Expiration date: 7/6/2013 EPA vendor ID: F12012 Cylinder contains: CO2,NO,SO2,BALN
2	3:40 PM	6.490	6.640	
3	4:02 PM	6.490	6.650	
Mean (%CO2)		6.490	6.640	
Mean Difference (%CO2)		0.150	Limit: N/A	
CEMS Accuracy		2%	Limit: 15% Passed	
Mid Gas				
1	3:29 PM	11.800	12.030	Allowable Reference Values: 10-14 %CO2 Cylinder ID: EB-0014601 Expiration date: 2/12/2021 EPA vendor ID: E12013 Cylinder contains: CO2,NO,SO2,BALN
2	3:51 PM	11.800	12.040	
3	4:13 PM	11.800	12.050	
Mean (%CO2)		11.800	12.040	
Mean Difference (%CO2)		0.240	Limit: N/A	
CEMS Accuracy		2%	Limit: 15% Passed	

Linearity Error (CEMS Accuracy) Determination (A): $A = (|C_m - C_a| / C_a) * 100$

C_a = Reference gas value

C_m = Mean of actual CEMS responses

Reactor B SO2 CGA

Indiantown CoGen

Test Information

Test Date: 3/7/2013
Facility: Indiantown CoGen
Unit: Reactor B

Analyzer Information

Range: Single
Instrument Span: 2000 ppm
Manufacturer: Thermo Environmental
Model: 43C
Serial Number: 43C-64047-342

Run Number	Time	Reference Gas	CEMS Response	Cylinder Information
------------	------	---------------	---------------	----------------------

Low Gas

1	3:18 PM	517.300	528.500
2	3:40 PM	517.300	528.500
3	4:02 PM	517.300	529.800

Mean (ppm) 517.300 528.933
Mean Difference (ppm) 11.633 *Limit: 5*
CEMS Accuracy 2% *Limit: 15% Passed*

Allowable Reference Values:
400-600 ppm (20-30% of span)
Cylinder ID: CC6414
Expiration date: 7/6/2013
EPA vendor ID: F12012
Cylinder contains:
CO2,NO,SO2,BALN

Mid Gas

1	3:29 PM	1140.000	1150.000
2	3:51 PM	1140.000	1153.300
3	4:13 PM	1140.000	1152.100

Mean (ppm) 1140.000 1151.800
Mean Difference (ppm) 11.800 *Limit: 5*
CEMS Accuracy 1% *Limit: 15% Passed*

Allowable Reference Values:
1000-1200 ppm (50-60% of span)
Cylinder ID: EB-0014601
Expiration date: 2/12/2021
EPA vendor ID: E12013
Cylinder contains:
CO2,NO,SO2,BALN

Linearity Error (CEMS Accuracy) Determination (A): $A = (|C_m - C_a| / C_a) * 100$

C_a = Reference gas value

C_m = Mean of actual CEMS responses

Section 4
Cylinder Gas Certification Sheets

DocNumber: 000005096

CERTIFICATE OF ANALYSIS / EPA PROTOCOL GAS

Customer & Order Information:

PD: WHSE DISTRIBUTION ORLAN
403 ZELL DR
ORLANDO FL 328240

Praxair Order Number: 02392292
Customer P. O. Number: 117278 91
Customer Reference Number:

Fill Date: 6/24/2010
Part Number: NI CD6.5S1E-AS
Lot Number: 917017503
Cylinder Style & Outlet: AS CGA660
Cylinder Pressure & Volume: 2000 psig 140 cu. ft.

Certified Concentration:

Expiration Date:	7/6/2013	NIST Traceable
Cylinder Number:	CC6414	Analytical Uncertainty:
517.3 ppm	SULFUR DIOXIDE	± 1 %
6.49 %	CARBON DIOXIDE	± 1 %
Balance	NITROGEN	

Certification Information: Certification Date: 7/6/2010 Term: 36 Months Expiration Date: 7/6/2013

This cylinder was certified according to the 1997 EPA Traceability Protocol, Document #EPA-600/R-97/121, using Procedure G1
Do Not Use this Standard if Pressure is less than 150 PSIG

Analytical Data:

(R=Reference Standard, Z=Zero Gas, C=Gas Candidate)

1. Component: SULFUR DIOXIDE

Requested Concentration: 500 ppm
Certified Concentration: 517.3 ppm
Instrument Used: SIEMENS ULTRAMAT 6E SN: J2-36
Analytical Method: NON-DISPERSIVE INFRARED
Last Multipoint Calibration: 6/23/2010

Reference Standard Type: GMIS
Ref. Std. Cylinder #: CC154814
Ref. Std. Conc: 509.2 PPM
Ref. Std. Traceable to SRM #: 1661a
SRM Sample #: 94-H-1
SRM Cylinder #: FF20869

First Analysis Data:	Date:	6/29/2010
Z: 0 R: 506.6 C: 514.5 Conc: 517.41		
R: 506.5 Z: 0 C: 514.3 Conc: 517.21		
Z: 0 C: 514.6 R: 505.9 Conc: 517.51		
UOM: PPM	Mean Test Assay:	517.38 PPM

Second Analysis Data:	Date:	7/6/2010
Z: 0 R: 510.8 C: 518.6 Conc: 516.98		
R: 510.9 Z: 0 C: 519.1 Conc: 517.47		
Z: 0 C: 519 R: 510.7 Conc: 517.37		
UOM: PPM	Mean Test Assay:	517.28 PPM

2. Component: CARBON DIOXIDE

Requested Concentration: 6.5 %
Certified Concentration: 6.49 %
Instrument Used: SIEMENS ULTRAMAT 5E SN: D2-412
Analytical Method: NON-DISPERSIVE INFRARED
Last Multipoint Calibration: 6/17/2010

Reference Standard Type: GMIS
Ref. Std. Cylinder #: CC167881
Ref. Std. Conc: 7.97 %
Ref. Std. Traceable to SRM #: 1674b
SRM Sample #: 7-G-27
SRM Cylinder #: CAL017236

First Analysis Data:	Date:	6/30/2010
Z: 0 R: 7.96 C: 6.48 Conc: 6.488		
R: 7.96 Z: 0 C: 6.48 Conc: 6.488		
Z: 0 C: 6.48 R: 7.96 Conc: 6.488		
UOM: %	Mean Test Assay:	6.488 %


Second Analysis Data:	Date:	
Z: 0 R: 0 C: 0 Conc: 0		
R: 0 Z: 0 C: 0 Conc: 0		
Z: 0 C: 0 R: 0 Conc: 0		
UOM: %	Mean Test Assay:	0 %

Analyzed by:

Michelle Kostik

Certified by:

Robin Morgan



LIQUID TECHNOLOGY CORPORATION

"INDUSTRY LEADER IN SPECIALTY GASES"

CGA MID
IN SVC.
3/6/13

Certificate of Analysis - EPA PROTOCOL GAS -

Customer Indiantown Cogen, LP (Indiantown, FL)
Date February 19, 2013
Delivery Receipt DR-45823
Gas Standard 1113 ppm SO₂, 12.0% CO₂/Nitrogen - EPA PROTOCOL
Final Analysis Date February 12, 2013
Expiration Date February 12, 2021

DO NOT USE BELOW 100 psig

Analytical Data:
EPA Protocol, Section No. 2.2, Procedure G-1.

Reported Concentrations

Sulfur Dioxide: 1140 ppm +/- 11.4 ppm

Carbon Dioxide: 11.8% +/- 0.07%

Nitrogen: Balance

** Total NOx for Reference Use Only **

Reference Standards

SRM/GMIS:	GMIS/GMIS	GMIS/GMIS
Cylinder Number:	CC-231460/CC-184275	EB-0026839/CC-185129
Concentration:	1015.69 ppm SO ₂ /1513.19 ppm SO ₂	6.847% CO ₂ /13.92% CO ₂
Expiration Date:	12/02/12 - 01/11/15	10/03/20 - 06/24/13

Certification Instrumentation

Component:	Sulfur Dioxide	Carbon Dioxide
Make/Model:	Nicolet 6700	Nicolet 6700
Serial Number:	APW1100563	APW1100563
Principal of Measurement:	FTIR	FTIR
Last Calibration:	January 30, 2013	January 30, 2013

Cylinder Data

Cylinder Number:	EB-0014601	Cylinder Volume:	126 Cubic Feet
Cylinder Outlet:	CGA 660	Cylinder Pressure:	1800 psig, 70°F

Analytical Uncertainty and NIST Traceability are in compliance with EPA-600/R-12/531.

Certified by:

David Scott

PGVP Vendor ID: E12013

"UNMATCHED EXCELLENCE"

Opacity Audits

OPACITY PERFORMANCE AUDIT

FOR

Indiantown Cogeneration

Indiantown, FL

Units: Main Emission unit E001

**DURAG
MODEL CEMOP-290 OPACITY CEMS**

First Quarter 2013

PREPARED BY:

Indiantown Cogeneration

TABLE OF CONTENTS

I. Introduction..... 1

II. Indiantown Cogen, Model CEMOP-290..... 3

A. CEMS Description.....3

B. Performance Audit Procedures.....4

C. Interpretation of Audit Results10

APPENDIX A MODEL CEMOP-290 AUDIT DATA FORMS

APPENDIX B FILTER DATA AND FIELD CERTIFICATION SHEETS

I. Introduction

Indiantown Cogen conducted an opacity performance audit on Durag Model CEMOP-290 opacity system. The testing was performed on March 8 2013 by Eric Rammel of Indiantown Cogen. The performance testing consists of:

1. Zero and Span Check
2. Transmissometer Dust Accumulation Check
3. Optical Alignment Check
4. Calibration Error Check

Procedures from U.S. EPA "Performance Audit Procedures for Opacity Monitors" were followed and are covered in Section 2. All raw data, calculated data and final summary are presented. The results indicate compliance for all specifications.

Facility Location: Indiantown Cogeneration, L.P. operates the Indiantown Cogeneration Plant, which is located in Martin County at 3303 SW Sliver Fox Lane, Indiantown, Florida.

The Indiantown Cogeneration Plant is a cogeneration facility which generates electricity for sale and exports steam to the Louis Dreyfus Citrus Processing Plant. The Facility includes one high-pressure pulverized coal main boiler (PC boiler) rated at 3,422 million British thermal units (MMBtu)/hour heat input, and has a nominal net electrical power output of approximately 330 megawatts (MW). It is permitted to fire natural gas, propane, or No. 2 fuel oil for startup, shutdown, or load changes.

Also included are two natural gas (or propane) fired identical auxiliary boilers used for supplying steam to the steam host during times when the PC boiler is offline, as well as during PC boiler startup and shutdown periods. The two identically sized packaged water-tube steam boilers have a combined rated maximum capacity of 350 MMBtu/hr.

Steam produced by the auxiliary boilers is not used to generate electricity. In addition, the facility has a variety of ancillary equipment needed to support operations as a coal-fired cogeneration plant.

Reviewed by: Nicholas Langer

Date: 3/20/2013

**PERFORMANCE AUDIT PROCEDURES FOR THE
DURAG INC. OPACITY MONITOR**

II. Durag Inc. Model CEMOP-290

The instrument is manufactured by the Durag Corporation and distributed by Monitoring Solutions, Inc.

A. CEMS Description

The Durag Inc. CEMOP-290 opacity monitoring system consists of four major components: the Transmissometer, the terminal control box, the air-purging system and the remote control unit and data acquisition equipment. The Transmissometer component consists of an optical transmitter/receiver (transceiver) unit mounted on one side of a stack or duct and a retro reflector unit mounted on the opposite side. The transceiver unit contains the light source, the photodiode detector, and the associated electronics. The transceiver uses a single-lamp, single detector system to determine effluent opacity. An LED light source is modulated electronically at 2 KHz to eliminate any ambient light interference. The modulated beam is configured to alternately produce reference and measurement signals so that the effects of variations in the optical and electronic components of the CEMS are minimized.

The display terminal control box mounted beside the transceiver unit provides on-stack readout of the opacity output from the transceiver and can be used as a diagnostic tool.

The air purging system serves a threefold purpose: 1) it provides an air window to keep exposed optical surfaces clean; 2) it protects the optical surfaces from condensation of stack gas moisture; and 3) it minimizes thermal conduction from the stack to the instrument. A standard installation has one air-purging system for each the transceiver and the retro reflector units.

The remote control unit communicates with the remote display unit via an RS 422 cable.

The opacity monitor measures the amount of light transmitted through the effluent from the transceiver to the retro reflector and back again. The control unit uses the effluent transmittance to calculate the optical density of the effluent at the monitor location, or the "path" optical density. In order to provide stack exit opacity data, the path optical density must be corrected. The correction factor is expressed as the ratio of the stack exit inside diameter to the inside diameter of the stack at the Transmissometer location. This ratio is called the "stack correction factor" (SCF) by Durag Inc. The following equations illustrate the relationship between this ratio, path optical density, and stack exit opacity.

$$L_x/L_1 = \text{stack correction factor}$$

where: L_x = stack exit inside diameter (ft)

L_1 = the stack inside diameter (or the duct width) at the monitor location (ft)

$$OP_x = 1 - \left(1 - \frac{\text{Opacity}}{100}\right)^{\text{corr. factor}}$$

OP_x = stack exit opacity (%)

B. Performance Audit Procedures

Preliminary Data

1. Obtain the stack exit inside diameter and the stack or duct inside diameter or width at the monitor location. Record these values in Blanks 1 and 2 of the Durag Inc. CEMOP-Performance Audit Data Sheet.

Note: Effluent handling system dimensions may be acquired from the following sources listed in descending order of reliability: 1) physical measurements, 2) construction drawings, 3) opacity monitor installation/certification documents, and 4) source personnel recollections.

2. Calculate the stack correction factor (divide the value in Blank 1 by the value in Blank 2). Record the result in Blank 3.

3. Record the source-cited stack correction factor in Blank 4.

Note: The stack correction factor is preset by the manufacturer using information supplied by the source. The value recorded in Blank 4 should be the value source personnel agree should be set inside the monitor.

4. Obtain the reference zero and span calibration values. Record these values in Blank 5 and Blank 6, respectively.

Note: The reference zero and span calibration values may not be the same as the values recorded during instrument installation and/or certification. The zero and span values recorded in Blank 5 and Blank 6 should be the reference values recorded during the most recent clear-path calibration of the CEMS.

Error Checks

The following steps describe the error codes for the Durag Inc. CEMOP-290 remote control unit. Unless otherwise noted, the audit can continue with error codes being present, provided the source has been informed of the fault conditions.

- Error code 100 = Transceiver blower fault
- Error code 200 = Transceiver filter plugged
- Error code 300 = Reflector blower fault
- Error code 400 = Reflector filter plugged

Note: If a FAULT occurred, an error code should be displaying on the stack mounted display and on the remote display. An explanation of the error codes can be found in the manual.

Instrument Range Check

5. Check the opacity CEMS measurement range by pressing the MOD button (the LED on the button will light up) and using the PLUS button to cycle through the displays.
6. Record the instrument range in Blank 11.

Reference Signal, Zero and Span Checks

7. Initiate the calibration cycle by pressing the arrow and plus buttons simultaneously and holding for approximately 5 seconds.

Note: The monitor will automatically cycle through the internal zero, external zero, span and stack taper ratio modes.

8. Record the internal zero milliamp value displayed on the control panel display in Blank 12.

Note: The internal zero checks the instrument reference signal. Since the instrument provides a full scale output of 4 to 20 milliamps, a value of 4 milliamps displayed on the control unit display represents a zero condition. After 1 1/2 minutes in the internal zero mode, the monitor will automatically switch to the external zero mode.

9. Record the external zero value (in milliamps) displayed on the control unit display in Blank 13. Record the external zero value (in percent opacity) displayed on the opacity data recorder in Blank 14.

Note: During the zero calibration check, the zero mirror is moved into the path of the measurement beam by a servomotor. The zero mechanism is designed to present the transceiver with a simulated clear-path condition. The daily zero check does not test the actual clear-path zero, nor does it provide a check of cross-stack parameters such as the optical alignment of the Transmissometer or drift in the reflectance of the retro reflector. The actual clear-path zero can only be checked during clear-stack or off-stack calibration of the CEMS. In addition to simulating the instrument clear-path zero, the zero mechanism allows the amount of dust on the transceiver optics (primary lens and zero mirror) to be quantified. After 1 1/2 minutes in the external zero mode, the CEMS will automatically enter the span mode.

10. Record in Blank 15 the span value (in milliamps) displayed on the control unit panel meter. Record the span value (in percent opacity) displayed on the data recorder in Blank 16. Go to the Transmissometer location.

Note: During the span calibration check, a servomotor moves a span filter into the path of the measurement beam while the zero mirror is in place. The span mechanism is designed to provide an indication of the upscale accuracy of the CEMS relative to the simulated clear-path zero. The monitor will output its stack correction factor (SCF) for 1 ½ minutes when the span portion of the calibration cycle is completed. The CEMS automatically returns to the measurement mode when the SCF portion of the calibration cycle is complete.

Retro reflector Dust Accumulation Check

11. Record the effluent opacity prior to cleaning the retroreflector optics in Blank 17.
12. Open the transceiver housing, inspect and clean the retroreflector optics, and close the housing.
13. Record the post-cleaning effluent opacity in Blank 18. Go to the transceiver location.

Transceiver Dust Accumulation Check

14. Record the pre-cleaning effluent opacity in Blank 19.
15. Open the transceiver, clean the optics (primary lens and zero mirror) and close the transceiver.
16. Record the post-cleaning effluent opacity in Blank 20.

Alignment Check

17. Determine the monitor alignment by looking through the alignment port of the side of the transceiver.
18. Observe whether the image is centered in the cross hairs and record this information (YES or NO) in Blank 21.

Calibration Error Check

The calibration error check is performed using three neutral density filters. Performing the calibration error check on-stack using the filters determines the linearity of the instrument response relative to the current clear-path zero setting. This calibration error check does not determine the accuracy of the actual instrument clear-path zero or the status of any cross-stack parameters. A true calibration check is performed by moving the on-stack components to a location with minimal ambient opacity, making sure that the proper path length and alignments are attained, and then placing the calibration filters in the measurement path.

19. Put the monitor in Filter Audit mode
20. Record the audit filter serial numbers and opacity values in Blanks 22, 23, and 24.
21. Remove the filters from their protective covers, inspect and if necessary, clean them.
22. Insert the low range neutral density filter into the filter slot located in front of the heated lens.
23. Wait approximately two minutes or until a clear value has been recorded and displayed on the data recorder.

Note: The audit data should be taken from a data recording/reporting device that presents instantaneous opacity (or opacity data with the shortest available integration period).
24. Record the CEMS response to the low range neutral density filter.
25. Remove the low range filter and insert the mid range neutral density filter.
26. Wait approximately two minutes and record the CEMS response to the mid range neutral density filter.
27. Remove the mid range filter and insert the high range filter.
28. Wait approximately two minutes and record the CEMS response to the high range neutral density filter.

29. Remove the high range filter, wait approximately two minutes, and record the zero value.
30. Repeat steps 29 through 36 until a total of five opacity readings are obtained for each neutral density filter.
31. If six-minute integrated opacity data are recorded, repeat steps 28 through 36 once more, changing the waiting periods to 13 minutes.
32. Record the six-minute integrated data.

Note: In order to acquire valid six-minute averaged opacity data, each filter must remain in for at least two consecutive six-minute periods; the first period will be invalid because it was in progress when the filter was inserted. A waiting period of 13 minutes is recommended.

33. When the calibration error check is complete, return the monitor to measuring mode. Close the transceiver head and the weather cover, and return to the CEMS control unit.

Final Control Unit Adjustment Reset

34. Return to the control unit location and reset the opacity instrument range to its original setting (Blank 11) if necessary.
35. Obtain a copy of the audit data from the data recorder.
36. Transcribe the calibration error response data from the data recorder to Blanks 25 through 50 of the audit form and complete the audit data calculations.

C. Interpretation of Audit Results

This section is designed to help the auditor interpret the CEMOP-290 performance audit results.

Stack Exit Correlation Error Check

The path length correction error in Blank 51 should be within $\pm 2\%$. This error exponentially affects the opacity readings, resulting in over - or underestimation of the stack exit opacity. The most common error in computing the optical path length correction factor is the use of the flange-to-flange distance in place of the stack/duct inside diameter at the monitor location. This error will result in underestimation of the stack exit opacity and can be identified by comparing the monitor optical path length to the flange-to-flange distance; the flange-to-flange distance should be greater by approximately two to four feet.

Error code analysis

Error codes are typically associated with parameters that the monitor manufacturer feels are critical to CEMS function, and to the collection of valid opacity data. The parameters associated with each of the error codes are found in the manufacturer's manual. With the exception of alarms that warn of elevated opacity levels (alarm or warning lamps), the error codes indicate that the CEMS is not functioning properly.

Control Panel Meter Error (Optional)

The accuracy of the control panel meter is important at sources using the meter during monitor adjustment and calibration. The accuracy of the control panel meter (Blank 52 and Blank 54) is determined by comparing the zero and span reference values to the panel meter output recorded during the CEMS calibration check.

Zero and Span Checks

The CEMOP-290 internal zero (Blank 12) should be set to indicate 0% opacity (equivalent to 3.7 - 4.3 mA). An external zero error (Blank 53) greater than 4% opacity is usually due to excessive dust accumulation on the optical surfaces, electronic drift or an electronic/mechanical offset of the data recorder. Excessive dust on the optical surfaces sufficient to cause a significant zero error would be indicated by the difference in the internal and external zero values and/or window alarm. Instrument span error (Blank 55) may be caused by the same problems that cause zero errors and may be identified in a similar fashion. A span error may also be caused by an inaccurately named span filter.

If the zero and span errors are due to a data recorder offset, both errors will be in the same direction and will be of the same magnitude.

The external zero displayed on the control unit panel meter also indicates the level of dust accumulation on the zero retroreflector and transceiver measurement window. The difference between the internal and external zero responses should equal the amount of dust found on the transceiver optics (Blank 75). To convert the zero responses to a value that represents lens dusting in percent opacity, use the following equation.

$$\text{Meter response in \% opacity} = 6.25 [(\text{Blank 13}) - (\text{Blank 12})]$$

Transmissometer Dust Accumulation Check

The results of the dust accumulation check (Blank 58) should not exceed 4%. A dust accumulation value of more than 4% opacity indicates that the air flow of the purge system and/or the cleaning frequency of the optical surfaces are inadequate. When determining the optical surface dust accumulation, the auditor should note whether the effluent opacity is relatively stable (within $\pm 2\%$ opacity) before and after cleaning the optical surfaces. If the effluent opacity is fluctuating by more than $\pm 2\%$, the dust accumulation analysis should be omitted.

Optical Alignment Check

When the transceiver and retroreflector are misaligned, a portion of the measurement beam that should be returned to the measurement detector is misdirected, resulting in a positive bias in the data reported by the CEMS. One of the most common causes of misalignment is vibration which may cause the on-stack components to shift slightly on the instrument mounting flanges. Another common cause of misalignment is thermal expansion and contraction of the structure on which the transmissometer is mounted. If the CEMS is being audited while the unit is off-line (cold stack), the results of the alignment analysis may not be representative of the alignment of the instrument when the stack or duct is at normal operating temperature.

Calibration Error

Calibration error results (Blanks 68, 69 and 70) in excess of $\pm 3\%$ are indicative of a non-linear or miss calibrated instrument. However, the absolute calibration accuracy of the monitor can be determined only when the instrument clear-path zero value is known. If the zero and span data are out-of-specification, the calibration error data will often be biased in the direction of the zero and span errors. Even if the zero and span data indicate that the CEMS is calibrated properly, the monitor may still be inaccurate due to error in the clear-path zero adjustment. The optimum calibration procedure involves using neutral density filters during clear-stack or off-stack CEMS calibration. This procedure would establish both the absolute calibration accuracy and linearity of the CEMS. If this procedure is impractical, and it is reasonable to assume that the clear-path zero is set correctly, the monitor's calibration can be set using either the neutral density filters or the internal zero and span values.

**APPENDIX A
MODEL CEMOP-290 AUDIT DATA FORMS**

AUDIT DATA SHEET
DURAG MODEL CEMOP-290 OPACITY CEMS

Corporation:	Indiantown Cogeneration LP	Plant / Site:	Indiantown Cogeneration LP
Process Unit/Stack Identification:	Main Stack		
Auditor:	Eric Rammel	Representing:	Indiantown Cogeneration LP
Attendees:	N/A	Representing:	N/A
Remote serial number:	412557		
Transceiver serial number:	413254	Flange to flange distance:	22' 0.75"
Reflector serial number:	413259		
Date:	3/8/2013		

Preliminary Data

1 Stack exit inside diameter (FT) = Lx	16
2 Stack (or duct) inside diameter at the transmissometer location (FT) = Lt	17.166
3 Calculated optical path length correction factor = Lx/Lt	0.93207503
4 Source-cited optical path length correction factor	0.9375
5 Source-cited zero automatic calibration value (% opacity)	0.00 %
6 Source-cited span automatic calibration value (% opacity)	80.00 %

[GO TO CONTROL UNIT / DATA RECORDER LOCATION]

[INSPECT DATA RECORDING SYSTEM AND MARK WITH "OPACITY AUDIT," AUDITOR'S NAME, AFFILIATION, DATE, SOURCE, PROCESS UNIT/STACK IDENTIFICATION, AND THE TIME OF DAY.]

Error codes

7 Blower [Loss of purge air blower power Error 100, 300]
 8 Filter [Inadequate purge air flow Error 200, 400]
 9 Window [Excessive dirt on transceiver window]
 10 Fault [Additional CEMS fault has occurred. Note fault code on panel meter and consult the instrument manual.]

ON	OFF
	X
	X
	X
	X

Instrument Range Check

11 Instrument range setting 100 %

Zero Check

12 Internal zero value (milliamps) 4.00 mA
 [Wait for two minutes for automatic change to external zero mode.]
 13 Panel meter zero calibration value (milliamps) 4.00 mA
 14 Opacity data recorder zero calibration value (% Op) 0.00 %
 [Wait two minutes for automatic change to external span mode.]

AUDIT DATA SHEET
 DURAG MODEL CEMOP-290 OPACITY CEMS
 (Continued)

Span Check

15 Panel meter span calibration value (milliamps) 11.50 mA
 16 Opacity data recorder span calibration value (% Op) 80.00 %
 [Go to transmissometer location.]

Retroreflector Dust Accumulation Check

17 Pre-cleaning effluent opacity (% Op) 0.70 %
 [Inspect and clean optical surface.]
 18 Post-cleaning effluent opacity (% Op) 0.00 %
 [Go to transceiver location.]

Transceiver Dust Accumulation Check

19 Pre-cleaning effluent opacity (% Op) 0.00 %
 [Inspect and clean optical surface.]
 20 Post-cleaning effluent opacity (% Op) 0.00 %

Optical Alignment Check (Optional)

[LOOK THROUGH ALIGNMENT SIGHT AND DETERMINE IF BEAM IMAGE IS CENTERED.]

21 Is the image centered?

YES	NO
X	

[Record audit filter data.]

Filter	Serial NO.	% Opacity	SCF%
22 LOW	RX55	8.50	7.99
23 MED	YA28	11.50	10.82
24 HIGH	RX56	23.20	21.92

AUDIT DATA SHEET
 DURAG MODEL CEMCP-290 OPACITY CEMS
 (Continued)

[Remove the audit filters from the protective covers, inspect, and clean each filter.]

[Insert a filter, wait approximately 2 minutes, and record the opacity value reported by the opacity data recorder. Repeat the process 5 times for each filter.]

<u>ZERO</u>	<u>LOW</u>	<u>MID</u>	<u>HIGH</u>	<u>ZERO</u>
<u>0.00</u>	<u>7.70</u>	<u>10.50</u>	<u>21.30</u>	<u>0.00</u>
	<u>7.70</u>	<u>10.60</u>	<u>21.40</u>	<u>0.00</u>
	<u>7.80</u>	<u>10.60</u>	<u>21.40</u>	<u>0.00</u>
	<u>7.80</u>	<u>10.60</u>	<u>21.40</u>	<u>0.00</u>
	<u>7.80</u>	<u>10.60</u>	<u>21.40</u>	<u>0.00</u>

[If six-minute integrated data are also available, allow 13 minutes each for an additional run of the ZERO, LOW, MID, HIGH, and ZERO readings.]

<u>ZERO</u>	<u>LOW</u>	<u>MID</u>	<u>HIGH</u>	<u>ZERO</u>
<u>0.00</u>	<u>8.30</u>	<u>10.70</u>	<u>21.40</u>	<u>0.00</u>

[Close the transceiver head and the weather cover.]

[Return to control unit location.]

[Obtain a copy of the audit data from the opacity data recorder, and ensure that the data can be clearly read and interpreted.]

[Read and transcribe final calibration error data.]

<u>ZERO</u>	<u>LOW</u>	<u>MID</u>	<u>HIGH</u>	<u>ZERO</u>
25 <u>0.00</u>	26 <u>7.70</u>	27 <u>10.50</u>	28 <u>21.30</u>	29 <u>0.00</u>
	30 <u>7.70</u>	31 <u>10.60</u>	32 <u>21.40</u>	33 <u>0.00</u>
	34 <u>7.80</u>	35 <u>10.60</u>	36 <u>21.40</u>	37 <u>0.00</u>
	38 <u>7.80</u>	39 <u>10.60</u>	40 <u>21.40</u>	41 <u>0.00</u>
	42 <u>7.80</u>	43 <u>10.60</u>	44 <u>21.40</u>	45 <u>0.00</u>

[Six-minute average data, if applicable.]

<u>ZERO</u>	<u>LOW</u>	<u>MID</u>	<u>HIGH</u>	<u>ZERO</u>
46 <u>0.00</u>	47 <u>8.30</u>	48 <u>10.70</u>	49 <u>21.40</u>	50 <u>0.00</u>

AUDIT DATA SHEET
 DURAG MODEL CEMOP-290 OPACITY CEMS
 (Continued)

Calculation of Audit Results

Stack exit correlation error (%):

	0.9375		0.932075032		
51	Blank 4	Blank 3	Blank 3	*100=	<u>0.58203125</u>
		0.932075			

Zero Error (% Op.):

		4		0	
52 Panel Meter	6.25 *	(Blank 13	---- 4.0)	--- Blank 5	=== <u>0.00 %</u>

		0		0	
53 Opacity Data Recorder		Blank 14	----	Blank 5	=== <u>0.00</u>

Span Error (% Op.):

		11.5		80	
54 Panel Meter	6.25 *	(Blank 15	---- 4.0)	--- Blank 6	=== <u>-33.13 %</u>

		80		80	
55 Opacity Data Recorder		Blank 16	----	Blank 6	=== <u>0.00</u>

Optical Surface Dust Accumulation (% OP):

		0.7		0	
56 Retroreflector		Blank 17	-----	Blank 18	=== <u>0.70 %</u>

		0		0	
57 Transceiver		Blank 19	-----	Blank 20	=== <u>0.00 %</u>

		0.7		0	
58 Total		Blank 56 +		Blank 57	=== <u>0.70 %</u>

Optical Path length Correction Factor and Zero Offset Correction of Audit Filters:

		0.9375			
59 LOW:	(1---	(1---	Blank 22))	*	(1- Blank 45) 100)
		8.5	Blank 4	0	<u>7.99058368 %</u>

		0.9375			
60 MID:	(1---	(1---	Blank 23))	*	(1- Blank 45) 100)
		11.5	Blank 4	0	<u>10.8216739 %</u>

		0.9375			
61 HIGH:	(1---	(1---	Blank 24))	*	(1- Blank 45) 100)
		23.2	Blank 4	0	<u>21.922456 %</u>

DURAG MODEL CEMCP-290 OPACITY CEMS
Performance Audit Data Summary

Auditor	Eric Rammel	Date	03/08/13
Source	Indiantown Cogeneration LP	Unit	Main Stack

PARAMETER	Blank No.	Audit Results	Specifications
Blower failure	7	0 X	OFF
Filter Block	8	0 X	OFF
Window	9	0 X	OFF
Fault	10	0 X	OFF
Stack Exit Correlation Error	51	0.58	+/- 2% Op
Internal Zero Error	Panel 52	0.00	+/- 4% Op
	Data 53	0.00	+/- 4% Op
Internal Span Error	Panel 54	-33.13	+/- 4% Op
	Data 55	0.00	+/- 4% Op
Optical Alignment Analysis	21	X	J
Optical Surface Dust Accumulation			
Retroreflector	56	0.70	<= 2% Op
Transceiver	57	0.00	<= 2% Op
Total	58	0.70	<= 4% Op
Calibration Error Analysis			
Mean Error			
Low	62	-0.23	
	71a	0.31	
Mid	63	-0.24	
	72a	-0.12	
High	64	-0.54	
	73a	-0.52	
Confidence Interval			
Low	65	0.07	
Mid	66	0.06	
High	67	0.06	
Calibration Error			
Low	68	0.30	<= 3% Op
Mid	69	0.30	<= 3% Op
High	70	0.60	<= 3% Op

n	Low range Difference		Delta L	Delta L ^2
1	7.7	7.990584	<u>-0.29</u>	<u>0.084</u>
	Blank 26 -	Blank59 =		
2	7.7	7.990584	<u>-0.29</u>	<u>0.084</u>
	Blank30 -	Blank59 =		
3	7.8	7.990584	<u>-0.19</u>	<u>0.036</u>
	Blank34 -	Blank59 =		
4	7.8	7.990584	<u>-0.19</u>	<u>0.036</u>
	Blank38 -	Blank59 =		
5	7.8	7.990584	<u>-0.19</u>	<u>0.036</u>
	Blank42 -	Blank59 =		
		Sum	-1.15	0.278
		DL		DL2

MEAN ERROR = ME low

$$\text{ME low} = \frac{\text{DL}}{n}$$

$$62 \text{ ME low} = \underline{-0.23058}$$

Confidence Interval = CI low

$$\text{CI low} = ((n * \text{DL}^2) - (\text{DL})^2)^{0.5} * 0.2776$$

$$65 \text{ CI low} = \underline{0.067998}$$

Calibration Error = CE low

$$\text{CE low} = \text{ME low} + \text{CI low}$$

$$68 \text{ CE low} = \underline{0.298582}$$

$$\text{E(6) low} = \frac{8.3}{\text{Blank 47}} - \frac{7.991}{\text{Blank 59}}$$

$$71 \text{ E(6) low} = \underline{0.309416}$$

n	Low range Difference		Delta M	Delta M ^2
1	10.5	10.82167	<u>-0.32</u>	<u>0.103</u>
	Blank 27 -	Blank60 =		
2	10.6	10.82167	<u>-0.22</u>	<u>0.049</u>
	Blank31 -	Blank60 =		
3	10.6	10.82167	<u>-0.22</u>	<u>0.049</u>
	Blank35 -	Blank60 =		
4	10.6	10.82167	<u>-0.22</u>	<u>0.049</u>
	Blank39 -	Blank60 =		
5	10.6	10.82167	<u>-0.22</u>	<u>0.049</u>
	Blank43 -	Blank60 =		
		Sum	-1.21	0.3
		DM		DM2

MEAN ERROR = ME mid

$$\text{ME mid} = \frac{\text{DM}}{n}$$

$$63 \text{ ME mid} = \underline{-0.24167}$$

Confidence Interval = CI mid

$$\text{CI mid} = ((n * \text{DM}^2) - (\text{DM})^2)^{0.5} * 0.2776$$

$$66 \text{ CI mid} = \underline{0.05552}$$

Calibration Error = CE mid

$$\text{CE mid} = \text{ME mid} + \text{CI mid}$$

$$69 \text{ CE mid} = \underline{0.297194}$$

$$\text{E(6) mid} = \frac{10.7}{\text{Blank 48}} - \frac{10.82}{\text{Blank 60}}$$

$$72 \text{ E(6) mid} = \underline{-0.12167}$$

n	Low range Difference		Delta H	Delta H ^2
1	21.3	21.92246	<u>-0.62</u>	<u>0.387451</u>
	Blank 28 -	Blank61 =		
2	21.4	21.92246	<u>-0.52</u>	<u>0.27296</u>
	Blank32 -	Blank61 =		
3	21.4	21.92246	<u>-0.52</u>	<u>0.27296</u>
	Blank35 -	Blank61 =		
4	21.4	21.92246	<u>-0.52</u>	<u>0.27296</u>
	Blank40 -	Blank61 =		
5	21.4	21.92246	<u>-0.52</u>	<u>0.27296</u>
	Blank44 -	Blank61 =		
		Sum	-2.71	1.479293
		DH		DH2

MEAN ERROR = ME high

$$\text{ME high} = \frac{\text{DH}}{n}$$

$$64 \text{ ME high} = \underline{-0.54246}$$

Confidence Interval = CI high

$$\text{CI high} = ((n * \text{DH}^2) - (\text{DH})^2)^{0.5} * 0.2776$$

$$67 \text{ CI high} = \underline{0.05552}$$

Calibration Error = CE high

$$\text{CE high} = \text{ME high} + \text{CI high}$$

$$70 \text{ CE high} = \underline{0.597976}$$

$$\text{E(6) high} = \frac{21.4}{\text{Blank 49}} - \frac{21.92}{\text{Blank 61}}$$

$$73 \text{ E(6) high} = \underline{-0.52246}$$

CeDAR 1-Minute Data

Indiantown CoGen

Data for 3/8/2013 10:30 AM thru 3/8/2013 1:30 PM

Timestamp	(Main Boiler) Opacity % 1-Min Avg	(Main Boiler) Opacity % 6-Min Avg
3/8 10:30	0.9	0.9
3/8 10:31	0.9	--
3/8 10:32	0.9	--
3/8 10:33	0.8	--
3/8 10:34	0.8	--
3/8 10:35	0.9	--
3/8 10:36	0.9	0.8
3/8 10:37	0.8	--
3/8 10:38	0.8	--
3/8 10:39	0.9	--
3/8 10:40	0.8	--
3/8 10:41	0.9	--
3/8 10:42	0.8	0.5 <14>
3/8 10:43	0.1	--
3/8 10:44	0.0 <14>	--
3/8 10:45	0.0 <14>	--
3/8 10:46	53.1 <14>	--
3/8 10:47	59.3 <14>	--
3/8 10:48	17.9 <14>	1.0 <15>
3/8 10:49	1.0 <14>	--
3/8 10:50	1.0	--
3/8 10:51	0.8 <15>	--
3/8 10:52	0.9 <15>	--
3/8 10:53	1.7 <15>	--
3/8 10:54	1.4 <15>	1.0 <15>
3/8 10:55	1.1 <15>	--
3/8 10:56	1.1 <15>	--
3/8 10:57	1.1 <15>	--
3/8 10:58	1.2 <15>	--
3/8 10:59	0.0 <15>	--
3/8 11:00	0.0 <15>	1.6 <15>
3/8 11:01	0.0 <15>	--
3/8 11:02	0.0 <15>	--
3/8 11:03	0.0 <15>	--
3/8 11:04	2.2 <15>	--
3/8 11:05	7.7 <15>	--
3/8 11:06	7.7 <15>	11.3 <15>
3/8 11:07	7.1 <15>	--
3/8 11:08	10.6 <15>	--
3/8 11:09	10.6 <15>	--
3/8 11:10	10.6 <15>	--
3/8 11:11	21.3 <15>	--
3/8 11:12	21.3 <15>	7.3 <15>
3/8 11:13	13.3 <15>	--
3/8 11:14	0.0 <15>	--
3/8 11:15	0.0 <15>	--

RUN 1 ZERO

RUN 1 LOW

RUN 1 MID

RUN 1 HIGH

RUN 2 ZERO

Timestamp	(Main Boiler) Opacity % 1-Min Avg	(Main Boiler) Opacity % 6-Min Avg	
3/8 11:16	1.6 <15>	--	
3/8 11:17	7.8 <15>	--	RUN 2 LOW
3/8 11:18	7.7 <15>	11.1 <15>	
3/8 11:19	7.5 <15>	--	
3/8 11:20	10.6 <15>	--	
3/8 11:21	10.6 <15>	--	RUN 2 MID
3/8 11:22	8.8 <15>	--	
3/8 11:23	21.4 <15>	--	
3/8 11:24	21.3 <15>	7.6 <15>	RUN 2 HIGH
3/8 11:25	14.2 <15>	--	
3/8 11:26	0.0 <15>	--	
3/8 11:27	0.0 <15>	--	RUN 3 ZERO
3/8 11:28	2.4 <15>	--	
3/8 11:29	7.8 <15>	--	
3/8 11:30	7.8 <15>	9.8 <15>	RUN 3 LOW
3/8 11:31	8.3 <15>	--	
3/8 11:32	10.6 <15>	--	
3/8 11:33	10.6 <15>	--	RUN 3 MID
3/8 11:34	10.6 <15>	--	
3/8 11:35	11.0 <15>	--	
3/8 11:36	21.4 <15>	13.1 <15>	RUN 3 HIGH
3/8 11:37	21.4 <15>	--	
3/8 11:38	21.4 <15>	--	
3/8 11:39	14.4 <15>	--	
3/8 11:40	0.0 <15>	--	
3/8 11:41	0.0 <15>	--	RUN 4 ZERO
3/8 11:42	0.0 <15>	4.5 <15>	
3/8 11:43	2.3 <15>	--	
3/8 11:44	7.8 <15>	--	
3/8 11:45	7.8 <15>	--	RUN 4 LOW
3/8 11:46	7.8 <13>	--	
3/8 11:47	6.9 <13>	--	
3/8 11:48	10.6 <13>	14.5 <13>	RUN 4 MID
3/8 11:49	10.6 <13>	--	
3/8 11:50	10.6 <13>	--	
3/8 11:51	12.2 <13>	--	
3/8 11:52	21.4 <13>	--	
3/8 11:53	21.4 <13>	--	RUN 4 HIGH
3/8 11:54	21.4 <13>	5.9 <13>	
3/8 11:55	12.3 <13>	--	
3/8 11:56	0.0 <13>	--	
3/8 11:57	0.0 <13>	--	RUN 5 ZERO
3/8 11:58	0.0 <13>	--	
3/8 11:59	2.0 <13>	--	
3/8 12:00	7.8 <13>	8.7 <13>	RUN 5 LOW
3/8 12:01	7.8 <13>	--	
3/8 12:02	7.8 <13>	--	
3/8 12:03	7.4 <13>	--	
3/8 12:04	10.6 <13>	--	
3/8 12:05	10.6 <13>	--	RUN 5 MID

Timestamp	(Main Boiler) Opacity % 1-Min Avg	(Main Boiler) Opacity % 6-Min Avg	
3/8 12:06	10.6 <13>	17.0 <13>	
3/8 12:07	12.1 <13>	--	
3/8 12:08	21.4 <13>	--	
3/8 12:09	21.4 <13>	--	RUN 5 HIGH
3/8 12:10	21.4 <13>	--	
3/8 12:11	15.0 <13>	--	
3/8 12:12	0.0 <13>	0.0 <13>	
3/8 12:13	0.0 <13>	--	
3/8 12:14	0.0 <13>	--	
3/8 12:15	0.0 <13>	--	
3/8 12:16	0.0 <13>	--	
3/8 12:17	0.0 <13>	--	
3/8 12:18	0.0 <13>	0.0 <13>	6 MIN ZERO
3/8 12:19	0.0 <13>	--	
3/8 12:20	0.0 <13>	--	
3/8 12:21	0.0 <13>	--	
3/8 12:22	0.0 <13>	--	
3/8 12:23	0.0 <13>	--	
3/8 12:24	0.0 <13>	4.3 <13>	
3/8 12:25	0.0 <13>	--	
3/8 12:26	2.4 <13>	--	
3/8 12:27	7.8 <13>	--	
3/8 12:28	7.8 <13>	--	
3/8 12:29	7.8 <13>	--	
3/8 12:30	7.8 <13>	7.8 <13>	
3/8 12:31	7.8 <13>	--	
3/8 12:32	7.8 <13>	--	
3/8 12:33	7.8 <13>	--	
3/8 12:34	7.8 <13>	--	
3/8 12:35	7.8 <13>	--	
3/8 12:36	7.8 <13>	8.3 <13>	6 MIN LOW
3/8 12:37	7.8 <13>	--	
3/8 12:38	7.8 <13>	--	
3/8 12:39	7.8 <13>	--	
3/8 12:40	7.6 <13>	--	
3/8 12:41	10.7 <13>	--	
3/8 12:42	10.7 <13>	10.7 <13>	
3/8 12:43	10.7 <13>	--	
3/8 12:44	10.7 <13>	--	
3/8 12:45	10.7 <13>	--	
3/8 12:46	10.7 <13>	--	
3/8 12:47	10.7 <13>	--	
3/8 12:48	10.7 <13>	10.7 <13>	6 MIN MID
3/8 12:49	10.7 <13>	--	
3/8 12:50	10.7 <13>	--	
3/8 12:51	10.7 <13>	--	
3/8 12:52	10.7 <13>	--	
3/8 12:53	10.7 <13>	--	
3/8 12:54	13.3 <13>	20.0 <13>	
3/8 12:55	21.4 <13>	--	

Timestamp	(Main Boiler) Opacity % 1-Min Avg	(Main Boiler) Opacity % 6-Min Avg
3/8 12:56	21.4 <13>	--
3/8 12:57	21.4 <13>	--
3/8 12:58	21.4 <13>	--
3/8 12:59	21.4 <13>	--
3/8 13:00	21.4 <13>	21.4 <13> 6 MIN HIGH
3/8 13:01	21.4 <13>	--
3/8 13:02	21.4 <13>	--
3/8 13:03	21.4 <13>	--
3/8 13:04	21.4 <13>	--
3/8 13:05	21.5 <13>	--
3/8 13:06	21.5 <13>	9.4 <13>
3/8 13:07	21.5 <13>	--
3/8 13:08	13.1 <13>	--
3/8 13:09	0.0 <13>	--
3/8 13:10	0.0 <13>	--
3/8 13:11	0.0 <13>	--
3/8 13:12	0.0 <13>	19.3 <29>
3/8 13:13	0.0 <13>	--
3/8 13:14	0.0 <29>	--
3/8 13:15	0.0 <29>	--
3/8 13:16	46.3 <29>	--
3/8 13:17	69.6 <29>	--
3/8 13:18	17.9 <29>	3.0 <29>
3/8 13:19	0.0 <29>	--
3/8 13:20	0.0 <13>	--
3/8 13:21	0.0 <13>	--
3/8 13:22	0.0 <13>	--
3/8 13:23	0.0 <13>	--
3/8 13:24	0.0 <13>	0.0 <13> 6 MIN ZERO
3/8 13:25	0.0 <13>	--
3/8 13:26	0.0 <13>	--
3/8 13:27	0.0 <13>	--
3/8 13:28	0.0 <13>	--
3/8 13:29	0.0 <13>	--
3/8 13:30	0.0 <13>	0.0 <13>
Average (all)	8.4	7.5
Total (all)	--	--
Minimum (all)	0.0	0.0
Maximum (all)	69.6	21.4
Average (valid values only)	0.8	0.9
Total (valid values only)	--	--
Count (valid values only)	15	2

<13> = Down
 <14> = Calibration
 <15> = Preventative Maintenance
 <29> = Process Down And Calibration

APPENDIX B
FILTER DATA AND FIELD CERTIFICATION SHEETS

CAL CHECK

11600 Black Horse Run, Raleigh, North Carolina 27613 Phone (919) 847-1898 FAX (919) 847-8005

REPORT OF CERTIFICATION OF NEUTRAL DENSITY AUDIT FILTERS

Report prepared for: **Indiantown Cogeneration, L.P.**

Date of Filter Certification: **December 4, 2012**

Date of Filter Expiration: **December 3, 2013**

Monitor Make/Model: **Durag D-R 290**

Audit Device/Filter Slot Angle of Incidence: **10 Degrees**

Path-Length Correction: **1.000 (Straight Stack)**

Table 1-1: Individual Filter Certification Data

Serial Number	Opacity Value (%)	Transmittance (%)	Optical Density	Previous Opacity (%)	Change in Opacity (%)	Accuracy (%)
RX55	8.5	91.5	0.0386	8.5	0.0	± 0.5
YA28	11.5	88.5	0.0531	11.4	0.1	± 0.5
RX56	23.2	76.8	0.1148	23.1	0.1	± 0.5