TABLE 297.310-1 CALIBRATION SCHEDULE
(version dated 10/07/96)

[Note: This table is referenced in Rule 62-297.310, F.A.C.]

ITEM	MINIMUM CALIBRATION FREQUENCY	REFERENCE INSTRUMENT	TOLERANCE
Liquid in glass thermometer	Annually	ASTM Hg in glass ref. thermometer or equivalent, or thermometric points	+/-2%
Bimetallic thermometer	Quarterly	Calib. liq. in glass thermometer	5 degrees F
Thermocouple	Annually	ASTM Hg in glass ref. thermometer, NBS calibrated reference and potentiometer	5 degrees F
Barometer	Monthly	Hg barometer or NOAA station	+/-1% scale
Pitot Tube	When required or when damaged	By construction or measurements in wind tunnel D greater than 16" and standard pitot tube	See EPA Method 2, Fig. 2-2 & 2-3
Probe Nozzles	Before each test or when nicked, dented, or corroded	Micrometer	+/-0.001" mean of at least three readings Max. deviation between readings .004”
Dry Gas Meter and Orifice Meter	1. Full Scale: When received, When 5% change observed, Annually	Spirometer or calibrated wet test or dry gas test meter	2%
	2. One Point: Semiannually		
	3. Check after each test series	Comparison check	5%

[electronic file name: 297310-1.doc]

Filename: Table 297 310-1.doc
Directory: C:\Documents and Settings\friday_b\Local Settings\Temporary Internet Files\OLK8E
Template: C:\Documents and Settings\friday_b\Application Data\Microsoft\Templates\Normal.dot
Title: TABLE 297.310-1 Calibration Schedule
Subject: TABLE 297.310-1 Calibration Schedule
Author: DARM
Keywords: TABLE 297.310-1 Calibration Schedule
Comments:
Creation Date: 9/5/2006 9:44:00 AM
Change Number: 2
Last Saved On: 9/5/2006 9:44:00 AM
Last Saved By: svec_e
Total Editing Time: 1 Minute
Last Printed On: 9/8/2006 10:36:00 AM
As of Last Complete Printing
Number of Pages: 1
Number of Words: 321 (approx.)
Number of Characters: 1,915 (approx.)